

LB 119399


7470A

INSTRUMENT ID: CV2

Linear

$\mu$  Abs.:

39737



A= 0.0000e+000

B= 2.5267e-004

C= -3.6640e-002

Rho= 0.9993165

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.021  | 0.021  | 230   | 0.000      | 230   |       |       |       |       |     |
| 0.2    | 0.200  | 0.164  | -0.036 | 793   | 0.0 %      | 793   |       |       |       |       | -18 |
| 2.5    | 2.500  | 2.410  | -0.090 | 9682  | 0.0 %      | 9682  |       |       |       |       | -4  |
| 5.0    | 5.000  | 5.270  | 0.270  | 21004 | 0.0 %      | 21004 |       |       |       |       | -5  |
| 7.5    | 7.500  | 7.331  | -0.169 | 29159 | 0.0 %      | 29159 |       |       |       |       | -2  |
| 10.0   | 10.000 | 10.004 | 0.004  | 39737 | 0.0 %      | 39737 |       |       |       |       | 0   |

## LB119399 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Mar 2022 10:05:23

| Sample ID       | Extended ID | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50          | 230    | -       | 0.0000   | 7470A  | PPB   | 29 Mar 2022 10:14:35 | S    | Std  |
| 0.2 - 1         | 50.2        | 793    | -       | 0.2000   | 7470A  | PPB   | 29 Mar 2022 10:16:52 | S    | Std  |
| 2.5 - 1         | 52.5        | 9682   | -       | 2.5000   | 7470A  | PPB   | 29 Mar 2022 10:21:42 | S    | Std  |
| 5.0 - 1         | 55          | 21004  | -       | 5.0000   | 7470A  | PPB   | 29 Mar 2022 10:23:59 | S    | Std  |
| 7.5 - 1         | 57.5        | 29159  | -       | 7.5000   | 7470A  | PPB   | 29 Mar 2022 10:26:15 | S    | Std  |
| 10.0 - 1        | 511         | 39737  | -       | 10.0000  | 7470A  | PPB   | 29 Mar 2022 10:38:42 | S    | Std  |
| ICV50 - 1       | ICV50       | 15052  | 3.7665  | -        | 7470A  | PPB   | 29 Mar 2022 10:41:28 | U    | SMPL |
| ICB50 - 1       | ICB50       | -8     | -0.0387 | -        | 7470A  | PPB   | 29 Mar 2022 10:43:42 | U    | SMPL |
| CCV40 - 1       | CCV40       | 21567  | 5.4127  | -        | 7470A  | PPB   | 29 Mar 2022 10:45:58 | U    | SMPL |
| CCB40 - 1       | CCB40       | -3     | -0.0374 | -        | 7470A  | PPB   | 29 Mar 2022 10:48:12 | U    | SMPL |
| CRA - 1         | CRA         | 760    | 0.1554  | -        | 7470A  | PPB   | 29 Mar 2022 10:50:28 | U    | SMPL |
| HighStd - 1     | HighStd     | 40305  | 10.1472 | -        | 7470A  | PPB   | 29 Mar 2022 10:52:43 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 28122  | 7.0689  | -        | 7470A  | PPB   | 29 Mar 2022 10:54:58 | U    | SMPL |
| PB143628BL - 1  | PBW         | 109    | -0.0091 | -        | 7470A  | PPB   | 29 Mar 2022 10:57:15 | U    | SMPL |
| PB143628BS - 1  | LCSW        | 15277  | 3.8234  | -        | 7470A  | PPB   | 29 Mar 2022 10:59:32 | U    | SMPL |
| N2094-02 - 1    | WC-2        | 28     | -0.0296 | -        | 7470A  | PPB   | 29 Mar 2022 11:01:48 | U    | SMPL |
| N2109-02 - 1    | TP-1        | 68     | -0.0195 | -        | 7470A  | PPB   | 29 Mar 2022 11:04:06 | U    | SMPL |
| N2109-02DUP - 1 | TP-1DUP     | 62     | -0.0210 | -        | 7470A  | PPB   | 29 Mar 2022 11:06:24 | U    | SMPL |
| N2109-02MS - 1  | TP-1MS      | 15080  | 3.7736  | -        | 7470A  | PPB   | 29 Mar 2022 11:11:26 | U    | SMPL |
| N2109-02MSD - 1 | TP-1MSD     | 15226  | 3.8105  | -        | 7470A  | PPB   | 29 Mar 2022 11:13:40 | U    | SMPL |
| CCV41 - 1       | CCV41       | 21818  | 5.4761  | -        | 7470A  | PPB   | 29 Mar 2022 11:15:56 | U    | SMPL |
| CCB41 - 1       | CCB41       | -8     | -0.0387 | -        | 7470A  | PPB   | 29 Mar 2022 11:18:12 | U    | SMPL |
| PB143589TB - 1  | PB143589TB  | 38     | -0.0270 | -        | 7470A  | PPB   | 29 Mar 2022 11:20:28 | U    | SMPL |
| N2109-02LX5 - 1 |             | 5      | -0.0354 | -        | 7470A  | PPB   | 29 Mar 2022 11:22:44 | U    | SMPL |
| N2109-02A - 1   |             | 15080  | 3.7736  | -        | 7470A  | PPB   | 29 Mar 2022 11:24:59 | U    | SMPL |
| CCV42 - 1       | CCV42       | 20114  | 5.0456  | -        | 7470A  | PPB   | 29 Mar 2022 11:27:15 | U    | SMPL |
| CCB42 - 1       | CCB42       | -78    | -0.0563 | -        | 7470A  | PPB   | 29 Mar 2022 11:29:31 | U    | SMPL |