

LB119402


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4644E



A= 0.0000e+000

B= 2.1245e-004

C= -1.2939e-001

Rho= 0.9987958

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	QSD
0.0	0.000	-0.014	-0.014	545	0.000	545					-
0.2	0.200	0.188	-0.012	1492	0.0 %	1492					-6
2.5	2.500	2.338	-0.162	11613	0.0 %	11613					-6
5.0	5.000	5.218	0.218	25171	0.0 %	25171					4
7.5	7.500	7.732	0.232	37003	0.0 %	37003					3
10.0	10.000	9.738	-0.262	46446	0.0 %	46446					-3

LB119402 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Mar 2022 13:19:57

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	545	-	0.0000	7471B	PPB	29 Mar 2022	13:23:32	S Std
0.2 - 1	50.2	1492	-	0.2000	7471B	PPB	29 Mar 2022	13:25:49	S Std
2.5 - 1	52.5	11613	-	2.5000	7471B	PPB	29 Mar 2022	13:30:38	S Std
5.0 - 1	55	25171	-	5.0000	7471B	PPB	29 Mar 2022	13:35:33	S Std
7.5 - 1	57.5	37003	-	7.5000	7471B	PPB	29 Mar 2022	13:40:29	S Std
10.0 - 1	7.10	46446	-	10.0000	7471B	PPB	29 Mar 2022	13:45:22	S Std
ICV51 - 1	ICV51	20294	4.1821	-	7471B	PPB	29 Mar 2022	13:48:09	U SMPL
ICB51 - 1	ICB51	388	-0.0470	-	7471B	PPB	29 Mar 2022	13:50:23	U SMPL
CCV43 - 1	CCV43	24780	5.1351	-	7471B	PPB	29 Mar 2022	13:55:13	U SMPL
CCB43 - 1	CCB43	415	-0.0412	-	7471B	PPB	29 Mar 2022	13:57:30	U SMPL
CRA - 1	CRA	1470	0.1829	-	7471B	PPB	29 Mar 2022	13:59:48	U SMPL
HighStd - 1	HighStd	46147	9.6746	-	7471B	PPB	29 Mar 2022	14:02:03	U SMPL
ChkStd - 1	ChkStd	34247	7.1464	-	7471B	PPB	29 Mar 2022	14:04:18	U SMPL
PB143702BL - 1	PBS	389	-0.0467	-	7471B	PPB	29 Mar 2022	14:06:38	U SMPL
PB143702BS - 1	LCSS	20200	4.1621	-	7471B	PPB	29 Mar 2022	14:08:58	U SMPL
N2095-01 - 1	SB-01-5.0-6.0	766	0.0333	-	7471B	PPB	29 Mar 2022	14:11:14	U SMPL
N2095-01DUP - 1	SB-01-5.0-6.0DUP	820	0.0448	-	7471B	PPB	29 Mar 2022	14:13:33	U SMPL
N2095-01MS - 1	SB-01-5.0-6.0MS	19521	4.0179	-	7471B	PPB	29 Mar 2022	14:15:50	U SMPL
N2095-01MSD - 1	SB-01-5.0-6.0MSD	18669	3.8369	-	7471B	PPB	29 Mar 2022	14:18:07	U SMPL
N2095-02 - 1	SB-02-5.0-6.0	566	-0.0091	-	7471B	PPB	29 Mar 2022	14:20:25	U SMPL
CCV44 - 1	CCV44	25513	5.2909	-	7471B	PPB	29 Mar 2022	14:22:42	U SMPL
CCB44 - 1	CCB44	401	-0.0442	-	7471B	PPB	29 Mar 2022	14:24:57	U SMPL
N2095-03 - 1	SB-03-5.0-6.0	1231	0.1321	-	7471B	PPB	29 Mar 2022	14:27:16	U SMPL
N2095-04 - 1	SB-04-4.0-5.0	31847	6.6365	-	7471B	PPB	29 Mar 2022	14:29:31	U SMPL
N2095-05 - 1	SB-05-5.0-6.0	1295	0.1457	-	7471B	PPB	29 Mar 2022	14:31:47	U SMPL
N2095-06 - 1	SB-06-4.0-5.0	23898	4.9478	-	7471B	PPB	29 Mar 2022	14:34:05	U SMPL
N2095-07 - 1	SB-07-2.0-3.0	16991	3.4804	-	7471B	PPB	29 Mar 2022	14:36:21	U SMPL
N2095-08 - 1	SB-08-9.0-10.0	16599	3.3971	-	7471B	PPB	29 Mar 2022	14:38:39	U SMPL
N2095-09 - 1	SB-09-3.0-4.0	11961	2.4117	-	7471B	PPB	29 Mar 2022	14:40:58	U SMPL
N2095-10 - 1	SB-10-4.0-5.0	18910	3.8881	-	7471B	PPB	29 Mar 2022	14:43:18	U SMPL
N2095-11 - 1	SB-11-4.0-5.0	45954	9.6336	-	7471B	PPB	29 Mar 2022	14:45:37	U SMPL
N2095-12 - 1	SB-12-5.0-6.0	95434	20.1456	-	7471B	PPB	29 Mar 2022	14:47:55	U SMPL
CCV45 - 1	CCV45	24954	5.1721	-	7471B	PPB	29 Mar 2022	14:50:16	U SMPL
CCB45 - 1	CCB45	390	-0.0465	-	7471B	PPB	29 Mar 2022	14:52:42	U SMPL
N2095-13 - 1	SB-13-9.0-10.0	59081	12.4224	-	7471B	PPB	29 Mar 2022	14:55:00	U SMPL
N2095-14 - 1	SB-14-4.0-5.0	9923	1.9788	-	7471B	PPB	29 Mar 2022	14:57:15	U SMPL
N2095-15 - 1	SB-15-5.0-6.0	510	-0.0210	-	7471B	PPB	29 Mar 2022	14:59:45	U SMPL
PB143703BL - 1	PBS	558	-0.0108	-	7471B	PPB	29 Mar 2022	15:02:03	U SMPL
PB143703BS - 1	LCSS	20006	4.1209	-	7471B	PPB	29 Mar 2022	15:04:19	U SMPL
N2107-01 - 1	SS1	47475	9.9567	-	7471B	PPB	29 Mar 2022	15:06:35	U SMPL
N2107-01DUP - 1	SS1DUP	48247	10.1207	-	7471B	PPB	29 Mar 2022	15:14:09	U SMPL
N2107-01MS - 1	SS1MS	69154	14.5624	-	7471B	PPB	29 Mar 2022	15:16:26	U SMPL
N2107-01MSD - 1	SS1MSD	66809	14.0642	-	7471B	PPB	29 Mar 2022	15:18:48	U SMPL
N2107-02 - 1	SS2	1117	0.1079	-	7471B	PPB	29 Mar 2022	15:21:11	U SMPL
CCV46 - 1	CCV46	24611	5.0992	-	7471B	PPB	29 Mar 2022	15:23:30	U SMPL
CCB46 - 1	CCB46	368	-0.0512	-	7471B	PPB	29 Mar 2022	15:25:45	U SMPL
N2107-03 - 1	SS5	13655	2.7716	-	7471B	PPB	29 Mar 2022	15:28:04	U SMPL
N2123-01 - 1	20220388-COMP-CS-1	290080	61.4983	-	7471B	PPB	29 Mar 2022	15:30:20	U SMPL
N2123-02 - 1	20220339-AMEDNDED-COMP-CS-1	279504	59.2514	-	7471B	PPB	29 Mar 2022	15:32:37	U SMPL
N2125-01 - 1	RO-124030	1133	0.1113	-	7471B	PPB	29 Mar 2022	15:35:13	U SMPL
N2125-03 - 1	164718	2931	0.4933	-	7471B	PPB	29 Mar 2022	15:37:50	U SMPL
N2140-01 - 1	MHC	664	0.0117	-	7471B	PPB	29 Mar 2022	15:40:06	U SMPL
N2142-01 - 1	CW-1	126662	26.7800	-	7471B	PPB	29 Mar 2022	15:42:23	U SMPL
CCV47 - 1	CCV47	25823	5.3567	-	7471B	PPB	29 Mar 2022	15:44:40	U SMPL
CCB47 - 1	CCB47	301	-0.0654	-	7471B	PPB	29 Mar 2022	15:47:09	U SMPL
N2095-01LX5 - 1		597	-0.0026	-	7471B	PPB	29 Mar 2022	15:49:27	U SMPL
N2095-01A - 1		19763	4.0693	-	7471B	PPB	29 Mar 2022	15:51:42	U SMPL
N2107-01LX5 - 1		10422	2.0848	-	7471B	PPB	29 Mar 2022	15:53:57	U SMPL
N2107-01A - 1		66375	13.9720	-	7471B	PPB	29 Mar 2022	16:01:32	U SMPL
N2095-12DLX5 - 1		20159	4.1534	-	7471B	PPB	29 Mar 2022	16:03:47	U SMPL
N2095-13DLX2 - 1		34898	7.2847	-	7471B	PPB	29 Mar 2022	16:09:00	U SMPL
N2123-01DLX10 - 1		32647	6.8065	-	7471B	PPB	29 Mar 2022	16:11:15	U SMPL
N2123-02DLX10 - 1		31820	6.6308	-	7471B	PPB	29 Mar 2022	16:13:34	U SMPL
N2142-01DLX5 - 1		24521	5.0801	-	7471B	PPB	29 Mar 2022	16:15:53	U SMPL
CCV48 - 1	CCV48	25281	5.2416	-	7471B	PPB	29 Mar 2022	16:18:13	U SMPL
CCB48 - 1	CCB48	384	-0.0478	-	7471B	PPB	29 Mar 2022	16:20:32	U SMPL
N2107-01DLX2 - 1		27728	5.7614	-	7471B	PPB	29 Mar 2022	16:22:51	U SMPL
N2107-01DUPDLX2 - 1		27757	5.7676	-	7471B	PPB	29 Mar 2022	16:25:07	U SMPL
N2107-01MSDLX2 - 1		38854	8.1252	-	7471B	PPB	29 Mar 2022	16:27:25	U SMPL
N2107-01MSDDLX2 - 1		37368	7.8095	-	7471B	PPB	29 Mar 2022	16:29:43	U SMPL
N2107-01LDLX10 - 1		4680	0.8649	-	7471B	PPB	29 Mar 2022	16:32:03	U SMPL
N2107-01ADLX2 - 1		37684	7.8766	-	7471B	PPB	29 Mar 2022	16:34:22	U SMPL
CCV49 - 1	CCV49	24891	5.1587	-	7471B	PPB	29 Mar 2022	16:36:39	U SMPL
CCB49 - 1	CCB49	383	-0.0480	-	7471B	PPB	29 Mar 2022	16:38:58	U SMPL