



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119404

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/29/2022	13:35
2	CAL2	1	Water	NA	NA	20.4	7.00	03/29/2022	13:36
3	CAL3	1	Water	NA	NA	20.4	10.03	03/29/2022	13:37
4	ICV	1	Water	NA	NA	20.5	7.02	03/29/2022	13:38
5	CCV1	1	Water	NA	NA	20.5	2.01	03/29/2022	13:40
6	N2152-01	1	Water	NA	NA	20.1	6.84	03/29/2022	13:45
7	N2152-03	1	Water	NA	NA	20.2	6.87	03/29/2022	13:47
8	N2152-03DUP	1	Water	NA	NA	20.3	6.88	03/29/2022	13:48
9	CCV2	1	Water	NA	NA	20.3	12.02	03/29/2022	13:50

WB 119404

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2152 WorkList ID : 158162 Department : Wet-Chemistry Date : 03-29-2022 13:26:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
03/28/2022	Water	N2152-01	pH <u>E</u>	Cool 4 deg C	TULL02	E31	PIT-11-DEWATERING-EFFL-UNTREATED	03/28/2022	9040C
03/28/2022	Water	N2152-03	pH <u>E</u>	Cool 4 deg C	TULL02	E31	PIT-11-DEWATERING-EFFL-TREATED	03/28/2022	9040C

Date/Time 03/24/22 13:30

Raw Sample Received by: 20 LWC

Raw Sample Relinquished by: CR SM

Date/Time 03/24/22 14:30

Raw Sample Received by: CR SM

Raw Sample Relinquished by: 20 LWC