

LB119421


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42797



A= 0.0000e+000

B= 2.3590e-004

C= 3.0317e-003

Rho= 0.9996180

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	48	0.000	48					-8
0.2	0.200	0.184	-0.016	766	0.0 %	766					6
2.5	2.500	2.643	0.143	11192	0.0 %	11192					-3
5.0	5.000	4.842	-0.158	20512	0.0 %	20512					-1
7.5	7.500	7.418	-0.082	31435	0.0 %	31435					1
10.0	10.000	10.099	0.099	42797	0.0 %	42797					

LB119421 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 30 Mar 2022 09:27:08					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	48	-	0.0000	7470A	PPB	30 Mar 2022 09:59:46	S	Std
0.2 - 1	50.2	766	-	0.2000	7470A	PPB	30 Mar 2022 10:02:04	S	Std
2.5 - 1	52.5	11192	-	2.5000	7470A	PPB	30 Mar 2022 10:04:21	S	Std
5.0 - 1	55	20512	-	5.0000	7470A	PPB	30 Mar 2022 10:06:38	S	Std
7.5 - 1	57.5	31435	-	7.5000	7470A	PPB	30 Mar 2022 10:11:30	S	Std
10.0 - 1	59.5	42797	-	10.0000	7470A	PPB	30 Mar 2022 10:16:18	S	Std
ICV52 - 1	ICV52	15344	3.6226	-	7470A	PPB	30 Mar 2022 10:19:18	U	SMPL
ICB52 - 1	ICB52	-33	-0.0048	-	7470A	PPB	30 Mar 2022 10:21:33	U	SMPL
CCV50 - 1	CCV50	22245	5.2505	-	7470A	PPB	30 Mar 2022 10:26:21	U	SMPL
CCB50 - 1	CCB50	-61	-0.0114	-	7470A	PPB	30 Mar 2022 10:28:36	U	SMPL
CRA - 1	CRA	823	0.1972	-	7470A	PPB	30 Mar 2022 10:30:52	U	SMPL
HighStd - 1	HighStd	41566	9.8083	-	7470A	PPB	30 Mar 2022 10:33:07	U	SMPL
ChkStd - 1	ChkStd	29607	6.9872	-	7470A	PPB	30 Mar 2022 10:35:22	U	SMPL
PB143720BL - 1	PBW	-90	-0.0182	-	7470A	PPB	30 Mar 2022 10:37:40	U	SMPL
PB143720BS - 1	LCSW	16600	3.9189	-	7470A	PPB	30 Mar 2022 10:39:57	U	SMPL
N2063-11 - 1	WC-1-(10-10.5)	-27	-0.0033	-	7470A	PPB	30 Mar 2022 10:42:13	U	SMPL
N2063-11DUP - 1	WC-1-(10-10.5)DUP	17	0.0070	-	7470A	PPB	30 Mar 2022 10:44:31	U	SMPL
N2063-11MS - 1	WC-1-(10-10.5)MS	15490	3.6570	-	7470A	PPB	30 Mar 2022 10:46:47	U	SMPL
N2063-11MSD - 1	WC-1-(10-10.5)MSD	14594	3.4457	-	7470A	PPB	30 Mar 2022 10:49:04	U	SMPL
N2063-12 - 1	WC-4-1-(10-10.5)	-26	-0.0031	-	7470A	PPB	30 Mar 2022 10:51:20	U	SMPL
CCV51 - 1	CCV51	21240	5.0134	-	7470A	PPB	30 Mar 2022 10:53:36	U	SMPL
CCB51 - 1	CCB51	-56	-0.0102	-	7470A	PPB	30 Mar 2022 10:55:52	U	SMPL
N2063-13 - 1	WC-4-2-(10-10.5)	5	0.0042	-	7470A	PPB	30 Mar 2022 10:58:08	U	SMPL
PB143721BL - 1	PBW	23	0.0085	-	7470A	PPB	30 Mar 2022 11:00:23	U	SMPL
PB143721BS - 1	LCSW	15710	3.7089	-	7470A	PPB	30 Mar 2022 11:02:39	U	SMPL
N2157-02 - 1	MHD	-33	-0.0048	-	7470A	PPB	30 Mar 2022 11:04:54	U	SMPL
N2157-05 - 1	TP2	12	0.0059	-	7470A	PPB	30 Mar 2022 11:07:10	U	SMPL
N2157-05DUP - 1	TP2DUP	5	0.0042	-	7470A	PPB	30 Mar 2022 11:09:27	U	SMPL
N2157-05MS - 1	TP2MS	15846	3.7410	-	7470A	PPB	30 Mar 2022 11:14:27	U	SMPL
N2157-05MSD - 1	TP2MSD	14962	3.5325	-	7470A	PPB	30 Mar 2022 11:16:43	U	SMPL
PB143738BL - 1	PBW	-27	-0.0033	-	7470A	PPB	30 Mar 2022 11:19:01	U	SMPL
PB143738BS - 1	LCSW	15242	3.5985	-	7470A	PPB	30 Mar 2022 11:23:55	U	SMPL
CCV52 - 1	CCV52	20437	4.8240	-	7470A	PPB	30 Mar 2022 11:26:11	U	SMPL
CCB52 - 1	CCB52	-68	-0.0130	-	7470A	PPB	30 Mar 2022 11:28:28	U	SMPL
N1645-07 - 1	LOD-MDL-WATER-01-QT1-2022	189	0.0476	-	7470A	PPB	30 Mar 2022 11:30:45	U	SMPL
N1645-07 - 1	LOD-MDL-WATER-01-QT1-2022	674	0.1620	-	7470A	PPB	30 Mar 2022 11:33:00	U	SMPL
N1645-08 - 1	LOQ-WATER-02-QT1-2022	805	0.1929	-	7470A	PPB	30 Mar 2022 11:35:15	U	SMPL
N2152-01 - 1	PIT-11-DEWATERING-EFFL-UNTREATED	50	0.0148	-	7470A	PPB	30 Mar 2022 11:37:31	U	SMPL
N2152-02 - 1	PIT-11-DEWATERING-EFFL-UNTREATED	13	0.0061	-	7470A	PPB	30 Mar 2022 11:39:47	U	SMPL
N2152-03 - 1	PIT-11-DEWATERING-EFFL-TREATED	177	0.0448	-	7470A	PPB	30 Mar 2022 11:42:04	U	SMPL
N2152-04 - 1	PIT-11-DEWATERING-EFFL-TREATED	50	0.0148	-	7470A	PPB	30 Mar 2022 11:44:20	U	SMPL
N2154-01 - 1	MW-5	37	0.0118	-	7470A	PPB	30 Mar 2022 11:46:37	U	SMPL
N2154-02 - 1	MW-4	89	0.0240	-	7470A	PPB	30 Mar 2022 11:48:54	U	SMPL
N2154-02DUP - 1	MW-4DUP	84	0.0228	-	7470A	PPB	30 Mar 2022 11:51:10	U	SMPL
CCV53 - 1	CCV53	22268	5.2559	-	7470A	PPB	30 Mar 2022 11:53:26	U	SMPL
CCB53 - 1	CCB53	-56	-0.0102	-	7470A	PPB	30 Mar 2022 11:55:41	U	SMPL
N2154-02MS - 1	MW-4MS	14673	3.4643	-	7470A	PPB	30 Mar 2022 11:57:57	U	SMPL
N2154-02MSD - 1	MW-4MSD	14522	3.4287	-	7470A	PPB	30 Mar 2022 12:00:13	U	SMPL
PB143720TB - 1	PB143684TB	-43	-0.0071	-	7470A	PPB	30 Mar 2022 12:02:29	U	SMPL
PB143700TB - 1	PB143700TB	12	0.0059	-	7470A	PPB	30 Mar 2022 12:04:45	U	SMPL
N2063-11LX5 - 1		-31	-0.0043	-	7470A	PPB	30 Mar 2022 12:07:01	U	SMPL
N2063-11A - 1		15467	3.6516	-	7470A	PPB	30 Mar 2022 12:09:17	U	SMPL
N2157-05LX5 - 1		-61	-0.0114	-	7470A	PPB	30 Mar 2022 12:11:33	U	SMPL
N2157-05A - 1		15420	3.6405	-	7470A	PPB	30 Mar 2022 12:13:50	U	SMPL
N2154-02LX5 - 1		-46	-0.0078	-	7470A	PPB	30 Mar 2022 12:16:07	U	SMPL
N2154-02A - 1		15074	3.5589	-	7470A	PPB	30 Mar 2022 12:18:23	U	SMPL
CCV54 - 1	CCV54	20755	4.8990	-	7470A	PPB	30 Mar 2022 12:20:39	U	SMPL
CCB54 - 1	CCB54	-106	-0.0220	-	7470A	PPB	30 Mar 2022 12:22:56	U	SMPL