



TOTAL VOLATILE SOLIDS 160.4

Run Number: LB119426

SUPERVISOR: apatel

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

TEMP1 IN: 104 °C 03/29/2022 16:00 TEMP1 OUT: 103 °C 03/30/2022 08:00
TEMP2 IN: 103 °C 03/30/2022 08:30 TEMP2 OUT: 103 °C 03/30/2022 10:00
TEMP3 IN: 540 °C 03/30/2022 10:30 TEMP3 OUT: 540 °C 03/30/2022 12:00
TEMP4 IN: 560 °C 03/30/2022 12:30 TEMP4 OUT: 560 °C 03/30/2022 14:00

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB119426BL	89.6317	89.6317	100	89.6318	89.6318	89.6318	89.6318	0.0000	0
2	N1873-33	75.5403	75.5403	100	75.5787	75.5787	75.5405	75.5405	0.0382	382

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$