



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119428

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/30/2022	12:20
2	CAL2	1	Water	NA	NA	20.4	7.00	03/30/2022	12:21
3	CAL3	1	Water	NA	NA	20.4	10.03	03/30/2022	12:25
4	ICV	1	Water	NA	NA	20.4	6.98	03/30/2022	12:27
5	CCV1	1	Water	NA	NA	20.3	2.01	03/30/2022	12:30
6	N2175-01	1	Water	NA	NA	20.6	7.57	03/30/2022	12:33
7	N2177-02	1	Water	NA	NA	20.4	6.52	03/30/2022	12:35
8	N2177-02DUP	1	Water	NA	NA	20.5	6.53	03/30/2022	12:36
9	CCV2	1	Water	NA	NA	20.5	12.03	03/30/2022	12:38

WORKLIST(Hardcopy Internal Chain)

2B 19628

WorkList Name : PH N2177 WorkList ID : 158207 Department : Wet-Chemistry Date : 03-30-2022 12:12:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/18/2022	Water	N2175-01	pH	Cool 4 deg C	MINO01	N22	023-OUTFALL	03/18/2022	9040C
03/29/2022	Water	N2177-02	pH	Cool 4 deg C	TETR06	E31	TT-039-IDWGW-20220329	03/29/2022	9040C

Date/Time 03/30/22 12:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 03/30/22 13:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]