

LB119432


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

49138



A= 0.0000e+000

B= 2.0440e-004

C= -2.9512e-002

Rho= 0.9999262

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.023	-0.023	30	0.000	30					-6
0.2	0.200	0.189	-0.011	1071	0.0 %	1071					-1
2.5	2.500	2.525	0.025	12498	0.0 %	12498					-1
5.0	5.000	5.070	0.070	24949	0.0 %	24949					0
7.5	7.500	7.425	-0.075	36469	0.0 %	36469					
10.0	10.000	10.014	0.014	49138	0.0 %	49138					

LB119432 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 30 Mar 2022 14:13:40					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-2	30	-	0.0000	7471B	PPB	30 Mar 2022 14:18:20	S	Std
0.2 - 1	50-2	1071	-	0.2000	7471B	PPB	30 Mar 2022 14:20:38	S	Std
2.5 - 1	52-5	12498	-	2.5000	7471B	PPB	30 Mar 2022 14:22:55	S	Std
5.0 - 1	55	24949	-	5.0000	7471B	PPB	30 Mar 2022 14:25:11	S	Std
7.5 - 1	57-5	36469	-	7.5000	7471B	PPB	30 Mar 2022 14:27:30	S	Std
10.0 - 1	510	49138	-	10.0000	7471B	PPB	30 Mar 2022 14:29:49	S	Std
ICV53 - 1	ICV53	19213	3.8976	-	7471B	PPB	30 Mar 2022 14:32:44	U	SMPL
ICB53 - 1	ICB53	-82	-0.0463	-	7471B	PPB	30 Mar 2022 14:34:58	U	SMPL
CCV55 - 1	CCV55	24670	5.0130	-	7471B	PPB	30 Mar 2022 14:37:16	U	SMPL
CCB55 - 1	CCB55	-126	-0.0553	-	7471B	PPB	30 Mar 2022 14:39:31	U	SMPL
CRA - 1	CRA	1037	0.1824	-	7471B	PPB	30 Mar 2022 14:41:49	U	SMPL
HighStd - 1	HighStd	48904	9.9664	-	7471B	PPB	30 Mar 2022 14:44:04	U	SMPL
ChkStd - 1	ChkStd	34057	6.9317	-	7471B	PPB	30 Mar 2022 14:48:57	U	SMPL
PB143722BL - 1	PBS	-175	-0.0653	-	7471B	PPB	30 Mar 2022 14:51:14	U	SMPL
PB143722BS - 1	LCSS	19486	3.9534	-	7471B	PPB	30 Mar 2022 14:53:35	U	SMPL
N2149-01 - 1	PILE-LINDEN-SWITCH	15248	3.0872	-	7471B	PPB	30 Mar 2022 14:55:51	U	SMPL
N2149-03 - 1	PILE-LINDEN-PROPANE	8540	1.7160	-	7471B	PPB	30 Mar 2022 14:58:11	U	SMPL
N2150-01 - 1	279	1996	0.3785	-	7471B	PPB	30 Mar 2022 15:00:29	U	SMPL
N2150-03 - 1	362	1979	0.3750	-	7471B	PPB	30 Mar 2022 15:02:48	U	SMPL
N2150-05 - 1	337	1655	0.3088	-	7471B	PPB	30 Mar 2022 15:05:04	U	SMPL
CCV56 - 1	CCV56	24218	4.9206	-	7471B	PPB	30 Mar 2022 15:10:07	U	SMPL
CCB56 - 1	CCB56	-119	-0.0538	-	7471B	PPB	30 Mar 2022 15:12:21	U	SMPL
N2151-01 - 1	TR-04-032822	1794	0.3372	-	7471B	PPB	30 Mar 2022 15:14:39	U	SMPL
N2155-01 - 1	22L076-A	2416	0.4643	-	7471B	PPB	30 Mar 2022 15:16:55	U	SMPL
N2155-02 - 1	22L076-B	5061	1.0049	-	7471B	PPB	30 Mar 2022 15:19:10	U	SMPL
N2155-03 - 1	22L076-C	2824	0.5477	-	7471B	PPB	30 Mar 2022 15:21:26	U	SMPL
N2155-04 - 1	22L076-D	2809	0.5446	-	7471B	PPB	30 Mar 2022 15:23:43	U	SMPL
N2155-05 - 1	22L076-E	2442	0.4696	-	7471B	PPB	30 Mar 2022 15:26:00	U	SMPL
N2155-06 - 1	22L076-F	3310	0.6470	-	7471B	PPB	30 Mar 2022 15:28:17	U	SMPL
N2155-07 - 1	22L076-G	3345	0.6542	-	7471B	PPB	30 Mar 2022 15:30:34	U	SMPL
N2155-08 - 1	22L076-H	1973	0.3738	-	7471B	PPB	30 Mar 2022 15:32:52	U	SMPL
N2155-09 - 1	22L076-I	2586	0.4991	-	7471B	PPB	30 Mar 2022 15:35:07	U	SMPL
CCV57 - 1	CCV57	25087	5.0982	-	7471B	PPB	30 Mar 2022 15:37:23	U	SMPL
CCB57 - 1	CCB57	-98	-0.0495	-	7471B	PPB	30 Mar 2022 15:39:39	U	SMPL
N2157-01 - 1	MHD	1178	0.2113	-	7471B	PPB	30 Mar 2022 15:41:57	U	SMPL
N2157-01DUP - 1	MHDDUP	1172	0.2100	-	7471B	PPB	30 Mar 2022 15:47:40	U	SMPL
N2157-01MS - 1	MHDMS	20131	4.0852	-	7471B	PPB	30 Mar 2022 15:49:55	U	SMPL
N2157-01MSD - 1	MHDMSD	19585	3.9736	-	7471B	PPB	30 Mar 2022 15:52:11	U	SMPL
N2157-04 - 1	TP2	734	0.1205	-	7471B	PPB	30 Mar 2022 15:54:30	U	SMPL
N2158-01 - 1	OR-02-032822	2888	0.5608	-	7471B	PPB	30 Mar 2022 15:56:49	U	SMPL
PB143749BL - 1	PB143749BL	15	-0.0264	-	7471B	PPB	30 Mar 2022 15:59:07	U	SMPL
PB143749BS - 1	PB143749BS	19905	4.0390	-	7471B	PPB	30 Mar 2022 16:01:25	U	SMPL
N1645-01 - 1	LOD-MDL-SOIL-01-QT1-2022	342	0.0404	-	7471B	PPB	30 Mar 2022 16:09:02	U	SMPL
N1645-01RE - 1	LOD-MDL-SOIL-01-QT1-2022	840	0.1422	-	7471B	PPB	30 Mar 2022 16:11:17	U	SMPL
CCV58 - 1	CCV58	24483	4.9748	-	7471B	PPB	30 Mar 2022 16:13:32	U	SMPL
CCB58 - 1	CCB58	-110	-0.0520	-	7471B	PPB	30 Mar 2022 16:15:48	U	SMPL
N1645-02 - 1	LOQ-SOIL-02-QT1-2022	977	0.1702	-	7471B	PPB	30 Mar 2022 16:18:06	U	SMPL
N2171-01 - 1	PSE-COMP-01	1710	0.3200	-	7471B	PPB	30 Mar 2022 16:20:22	U	SMPL
N2171-01DUP - 1	PSE-COMP-01DUP	1581	0.2936	-	7471B	PPB	30 Mar 2022 16:22:38	U	SMPL
N2171-01MS - 1	PSE-COMP-01MS	22033	4.4740	-	7471B	PPB	30 Mar 2022 16:24:53	U	SMPL
N2171-01MSD - 1	PSE-COMP-01MSD	21400	4.3446	-	7471B	PPB	30 Mar 2022 16:27:09	U	SMPL
N2176-01 - 1	TT-156-IDWSO-20220329-1	-25	-0.0346	-	7471B	PPB	30 Mar 2022 16:29:33	U	SMPL
N2157-01LX5 - 1		248	0.0212	-	7471B	PPB	30 Mar 2022 16:31:53	U	SMPL
N2157-01A - 1		19800	4.0176	-	7471B	PPB	30 Mar 2022 16:34:10	U	SMPL
N2171-01LX5 - 1		262	0.0240	-	7471B	PPB	30 Mar 2022 16:36:27	U	SMPL
N2171-01A - 1		21959	4.4589	-	7471B	PPB	30 Mar 2022 16:38:45	U	SMPL
CCV59 - 1	CCV59	24015	4.8791	-	7471B	PPB	30 Mar 2022 16:41:01	U	SMPL
CCB59 - 1	CCB59	-123	-0.0547	-	7471B	PPB	30 Mar 2022 16:43:19	U	SMPL