



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119441

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	03/30/2022	17:15
2	CAL2	1	Water	NA	NA	20.9	7.00	03/30/2022	17:16
3	CAL3	1	Water	NA	NA	20.8	10.03	03/30/2022	17:19
4	ICV	1	Water	NA	NA	20.8	7.01	03/30/2022	17:20
5	CCV1	1	Water	NA	NA	20.8	2.01	03/30/2022	17:22
6	N2202-02	1	Solid	20.03	20	20.6	7.69	03/30/2022	17:25
7	N2202-05	1	Solid	20.04	20	21.0	7.97	03/30/2022	17:27
8	N2202-05DUP	1	Solid	20.03	20	21.1	7.98	03/30/2022	17:28
9	CCV2	1	Water	NA	NA	20.8	12.03	03/30/2022	17:30

VB 719647

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N2202 WorkList ID : 158221 Department : Wet-Chemistry Date : 03-30-2022 17:09:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
03/30/2022	Solid	N2202-02	Corrosivity	Cool 4 deg C	PSEG03	O11	MHE	03/30/2022	9045D
03/30/2022	Solid	N2202-05	Corrosivity	Cool 4 deg C	PSEG03	O11	TP3	03/30/2022	9045D

Date/Time 03/30/22 17:10
Raw Sample Received by: JB (WC)
Raw Sample Relinquished by: CD SM

Date/Time 03/30/22 18:20
Raw Sample Received by: CD SM
Raw Sample Relinquished by: JB (WC)