

LB119443


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41092



A= 0.0000e+000

B= 2.4247e-004

C= -2.7708e-002

Rho= 0.9997891

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/oD
0.0	0.000	-0.028	-0.028	-3	0.000	-3					-
0.2	0.200	0.168	-0.032	807	0.0 %	807					-16
2.5	2.500	2.481	-0.019	10347	0.0 %	10347					-1
5.0	5.000	5.166	0.166	21421	0.0 %	21421					3
7.5	7.500	7.477	-0.023	30952	0.0 %	30952					0
10.0	10.000	9.936	-0.064	41092	0.0 %	41092					-1

LB119443 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Mar 2022 09:04:28

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7470A	PPB	31 Mar 2022 09:07:48	S	Std
0.2 - 1	50.2	807	-	0.2000	7470A	PPB	31 Mar 2022 09:10:05	S	Std
2.5 - 1	52.5	10347	-	2.5000	7470A	PPB	31 Mar 2022 09:12:22	S	Std
5.0 - 1	55	21421	-	5.0000	7470A	PPB	31 Mar 2022 09:14:40	S	Std
7.5 - 1	57.5	30952	-	7.5000	7470A	PPB	31 Mar 2022 09:16:57	S	Std
10.0 - 1	510	41092	-	10.0000	7470A	PPB	31 Mar 2022 09:26:49	S	Std
ICV54 - 1	ICV54	16407	3.9505	-7470A	PPB	PPB	31 Mar 2022 09:32:08	U	SMPL
ICB54 - 1	ICB54	-34	-0.0360	-7470A	PPB	PPB	31 Mar 2022 09:34:23	U	SMPL
CCV60 - 1	CCV60	18960	4.5695	-7470A	PPB	PPB	31 Mar 2022 09:36:38	U	SMPL
CCB60 - 1	CCB60	-17	-0.0318	-7470A	PPB	PPB	31 Mar 2022 09:38:53	U	SMPL
CRA - 1	CRA	843	0.1767	-7470A	PPB	PPB	31 Mar 2022 09:41:09	U	SMPL
HighStd - 1	HighStd	41515	10.0384	-7470A	PPB	PPB	31 Mar 2022 09:43:25	U	SMPL
ChkStd - 1	ChkStd	27233	6.5755	-7470A	PPB	PPB	31 Mar 2022 09:48:18	U	SMPL
PB143750BL - 1	PBW	-73	-0.0454	-7470A	PPB	PPB	31 Mar 2022 09:50:34	U	SMPL
PB143750BS - 1	LCSW	16233	3.9083	-7470A	PPB	PPB	31 Mar 2022 09:55:22	U	SMPL
N1645-07 - 1	LOD-MDL-WATER-01-QT1-2022	339	0.0545	-7470A	PPB	PPB	31 Mar 2022 09:57:38	U	SMPL
N1645-07RE - 1	LOD-MDL-WATER-01-QT1-2022RE	740	0.1517	-7470A	PPB	PPB	31 Mar 2022 09:59:55	U	SMPL
N1645-09 - 1	MDL-WATER-03-QT1-2022	291	0.0429	-7470A	PPB	PPB	31 Mar 2022 10:12:42	U	SMPL
N2177-02 - 1	TT-039-IDWGW-20220329	126	0.0028	-7470A	PPB	PPB	31 Mar 2022 10:14:58	U	SMPL
N2177-02DUP - 1	TT-039-IDWGW-20220329DUP	119	0.0011	-7470A	PPB	PPB	31 Mar 2022 10:17:14	U	SMPL
CCV61 - 1	CCV61	21133	5.0964	-7470A	PPB	PPB	31 Mar 2022 10:19:29	U	SMPL
CCB61 - 1	CCB61	-21	-0.0328	-7470A	PPB	PPB	31 Mar 2022 10:21:44	U	SMPL
N2177-02MS - 1	TT-039-IDWGW-20220329MS	15192	3.6559	-7470A	PPB	PPB	31 Mar 2022 10:24:01	U	SMPL
N2177-02MSD - 1	TT-039-IDWGW-20220329MSD	16126	3.8824	-7470A	PPB	PPB	31 Mar 2022 10:26:16	U	SMPL
N2177-02LX5 - 1		-17	-0.0318	-7470A	PPB	PPB	31 Mar 2022 10:28:32	U	SMPL
N2177-02A - 1		18351	4.4219	-7470A	PPB	PPB	31 Mar 2022 10:30:48	U	SMPL
CCV62 - 1	CCV62	21367	5.1531	-7470A	PPB	PPB	31 Mar 2022 10:33:04	U	SMPL
CCB62 - 1	CCB62	-46	-0.0389	-7470A	PPB	PPB	31 Mar 2022 10:35:21	U	SMPL