

LB119453


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37383



A= 0.0000e+000

B= 2.6694e-004

C= -9.5221e-003

Rho= 0.9998145

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	70D
0.0	0.000	-0.003	-0.003	25	0.000	25					-
0.05	0.050	0.042	-0.008	192	0.0 %	192	0				-16
0.2	0.200	0.203	0.003	798	0.0 %	798					2
2.5	2.500	2.422	-0.078	9109	0.0 %	9109					-3
5.0	5.000	5.163	0.163	19376	0.0 %	19376					3
7.5	7.500	7.453	-0.047	27956	0.0 %	27956					-1
10.0	10.000	9.970	-0.030	37383	0.0 %	37383					0

LB119453 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 31 Mar 2022 12:42:52

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	245.1	PPB	31 Mar 2022 13:08:31	S	Std
0.05 - 1	50.5	192	-	0.0500	245.1	PPB	31 Mar 2022 13:10:48	S	Std
0.2 - 1	50.2	798	-	0.2000	245.1	PPB	31 Mar 2022 13:13:05	S	Std
2.5 - 1	52.5	9109	-	2.5000	245.1	PPB	31 Mar 2022 13:15:22	S	Std
5.0 - 1	55	19376	-	5.0000	245.1	PPB	31 Mar 2022 13:17:38	S	Std
7.5 - 1	57.5	27956	-	7.5000	245.1	PPB	31 Mar 2022 13:19:54	S	Std
10.0 - 1	510	37383	-	10.0000	245.1	PPB	31 Mar 2022 13:22:10	S	Std
ICV55 - 1	ICV55	14455	3.8491	-245.1		PPB	31 Mar 2022 13:25:03	U	SMPL
ICB55 - 1	ICB55	-59	-0.0253	-245.1		PPB	31 Mar 2022 13:27:18	U	SMPL
CCV63 - 1	CCV63	19729	5.2570	-245.1		PPB	31 Mar 2022 13:29:33	U	SMPL
CCB63 - 1	CCB63	-57	-0.0247	-245.1		PPB	31 Mar 2022 13:31:48	U	SMPL
CRA - 1	CRA	187	0.0404	-245.1		PPB	31 Mar 2022 13:34:04	U	SMPL
HighStd - 1	HighStd	37753	10.0684	-245.1		PPB	31 Mar 2022 13:36:19	U	SMPL
ChkStd - 1	ChkStd	25611	6.8272	-245.1		PPB	31 Mar 2022 13:38:34	U	SMPL
PB143767BL - 1	PBW	-98	-0.0357	-245.1		PPB	31 Mar 2022 13:40:51	U	SMPL
PB143767BS - 1	LCSW	14473	3.8540	-245.1		PPB	31 Mar 2022 13:43:09	U	SMPL
N1645-10 - 1	LOD-MDL-WATER-04-QT1-2022	212	0.0471	-245.1		PPB	31 Mar 2022 13:48:00	U	SMPL
N1645-10RE - 1	LOD-MDL-WATER-04-QT1-2022RE	658	0.1661	-245.1		PPB	31 Mar 2022 13:50:16	U	SMPL
N1645-11 - 1	LOQ-WATER-05-QT1-2022	745	0.1894	-245.1		PPB	31 Mar 2022 13:52:33	U	SMPL
N2086-01 - 1	EFFLUENT-DAY-1-METALS	109	0.0196	-245.1		PPB	31 Mar 2022 14:02:12	U	SMPL
N2086-01DUP - 1	EFFLUENT-DAY-1-METALSDUP	97	0.0164	-245.1		PPB	31 Mar 2022 14:04:27	U	SMPL
CCV64 - 1	CCV64	18171	4.8411	-245.1		PPB	31 Mar 2022 14:09:18	U	SMPL
CCB64 - 1	CCB64	-164	-0.0533	-245.1		PPB	31 Mar 2022 14:11:32	U	SMPL
N2086-01MS - 1	EFFLUENT-DAY-1-MET	14524	3.8676	-245.1		PPB	31 Mar 2022 14:13:49	U	SMPL
N2086-01MSD - 1	EFFLUENT-DAY-1-METALMSD	16278	4.3358	-245.1		PPB	31 Mar 2022 14:23:49	U	SMPL
N2086-01LX5 - 1		84	0.0129	-245.1		PPB	31 Mar 2022 14:31:40	U	SMPL
N2086-01A - 1		12672	3.3732	-245.1		PPB	31 Mar 2022 14:36:33	U	SMPL
CCV65 - 1	CCV65	17116	4.5595	-245.1		PPB	31 Mar 2022 14:41:26	U	SMPL
CCB65 - 1	CCB65	-197	-0.0621	-245.1		PPB	31 Mar 2022 14:43:42	U	SMPL