



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119457

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP96087
calibration std. phosphate 0.5 ppm	WP96086
calibration std. phosphate 0.3 ppm	WP96085
calibration std. phosphate 0.1 ppm	WP96084
calibration std. phosphate 0.05 ppm	WP96083
calibration std. 0 ppm	WP96082
phosphate CCV std.	WP96090
Combined reagent	WP96106
Phenolphthalein indicator	WP93767
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP96091

Intercept: -0.0013

Slope: 0.6534

Regression: 0.999955

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		03/31/2022	13:30
2	CAL2	0.05	1	50	50	0.033	0.052	-4	03/31/2022	13:30
3	CAL3	0.10	1	50	50	0.064	0.1	0	03/31/2022	13:31
4	CAL4	0.30	1	50	50	0.190	0.293	2.3	03/31/2022	13:31
5	CAL5	0.50	1	50	50	0.326	0.501	-0.2	03/31/2022	13:32
6	CAL6	1.00	1	50	50	0.653	1.001	-0.1	03/31/2022	13:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119457

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.495	03/31/2022	13:33
2	ICB		1	50	50	0.001	0.004	03/31/2022	13:33
3	CCV1	0.50	1	50	50	0.328	0.504	03/31/2022	13:34
4	CCB1		1	50	50	0.001	0.004	03/31/2022	13:34
5	PB143777BL		1	50	50	0.002	0.005	03/31/2022	13:35
6	PB143777BS	0.50	1	50	50	0.322	0.495	03/31/2022	13:35
7	N1873-13		1	50	50	0.994	1.523	03/31/2022	13:36
8	N2168-01		1	50	50	1.746	2.674	03/31/2022	13:36
9	N2185-01		1	50	50	0.975	1.494	03/31/2022	13:37
10	N2191-01		1	50	50	0.630	0.966	03/31/2022	13:37
11	N2191-01DUP		1	50	50	0.631	0.968	03/31/2022	13:38
12	N2191-01MS	0.50	1	50	50	0.673	1.032	03/31/2022	13:38
13	N2191-01MSD	0.50	1	50	50	0.675	1.035	03/31/2022	13:39
14	N2168-01		20	50	50	0.417	0.640	03/31/2022	13:39
15	CCV2	0.50	1	50	50	0.326	0.501	03/31/2022	13:40
16	CCB2		1	50	50	0.002	0.005	03/31/2022	13:40
17	N2185-01		5	50	50	0.547	0.839	03/31/2022	13:41
18	N2191-01MS	0.50	2	50	50	0.455	0.698	03/31/2022	13:41
19	N2191-01MSD	0.50	2	50	50	0.458	0.703	03/31/2022	13:42
20	CCV3	0.50	1	50	50	0.329	0.506	03/31/2022	13:42
21	CCB3		1	50	50	0.001	0.004	03/31/2022	13:43
22	N1873-13		10	50	50	0.237	0.365	03/31/2022	13:44
23	CCV4	0.50	1	50	50	0.325	0.499	03/31/2022	13:44
24	CCB4		1	50	50	0.001	0.004	03/31/2022	13:45