



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119458

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	03/31/2022	13:46
2	CAL2	1	Water	NA	NA	20.6	7.00	03/31/2022	13:47
3	CAL3	1	Water	NA	NA	20.6	10.03	03/31/2022	13:48
4	ICV	1	Water	NA	NA	20.7	6.98	03/31/2022	13:50
5	CCV1	1	Water	NA	NA	20.6	2.01	03/31/2022	14:00
6	N2215-03	1	Water	NA	NA	20.4	9.56	03/31/2022	14:10
7	N2215-03DUP	1	Water	NA	NA	20.5	9.57	03/31/2022	14:11
8	N2215-04	1	Water	NA	NA	20.7	3.33	03/31/2022	14:12
9	N2215-05	1	Water	NA	NA	20.6	7.48	03/31/2022	14:15
10	CCV2	1	Water	NA	NA	20.6	12.02	03/31/2022	14:20

WORKLIST(Hardcopy Internal Chain)

13 19458

WorkList Name : PH N2215 W WorkList ID : 158257 Department : Wet-Chemistry Date : 03-31-2022 13:43:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/31/2022	Water	N2215-03	pH <i>A</i>	Cool 4 deg C	PSEG03	O11	GW-1	03/31/2022	9040C
03/31/2022	Water	N2215-04	pH <i>A</i>	Cool 4 deg C	PSEG03	O11	GW-2	03/31/2022	9040C
03/31/2022	Water	N2215-05	pH <i>A</i>	Cool 4 deg C	PSEG03	O11	GW-3	03/31/2022	9040C

Date/Time 03/31/2022 13:45
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: At Sm

Date/Time 03/31/22 15:00
 Raw Sample Received by: At Sm
 Raw Sample Relinquished by: JP (WC)