



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB119460

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP96120
Turbidity Calibration std, 20NTU	WP96115
Turbidity Calibration std, 40NTU	WP96114
Turbidity Calibration std, 1NTU	WP96118
Turbidity Calibration std, 5NTU	WP96117
Turbidity Calibration - CCV std, 10 NTU	WP96116
10 NTU Standard 500 ml	W2752

Intercept: 0.1608

Slope: 0.9995

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.155	-0.01		03/31/2022	14:30
2	CAL2	1	1	1.160	1.00	0	03/31/2022	14:33
3	CAL3	5	1	5.230	5.07	-1.4	03/31/2022	14:36
4	CAL4	10	1	9.980	9.82	1.8	03/31/2022	14:39
5	CAL5	20	1	20.300	20.15	-0.7	03/31/2022	14:42
6	CAL6	40	1	40.100	39.96	0.1	03/31/2022	14:45



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB119460

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.500	10.344	03/31/2022	14:48
2	ICB		1	0.137	-0.024	03/31/2022	14:51
3	CCV1	10	1	10.300	10.144	03/31/2022	14:54
4	CCB1		1	0.142	-0.019	03/31/2022	14:57
5	LB119460BL		1	0.139	-0.022	03/31/2022	15:00
6	N1873-35		1	12.400	12.245	03/31/2022	15:03
7	CCV2	10	1	10.200	10.044	03/31/2022	15:05
8	CCB2		1	0.135	-0.026	03/31/2022	15:08