


LB119462

7471B

INSTRUMENT ID: CV1

 μ Abs.:

41693



Linear

A= 0.0000e+000

B= 2.4066e-004

C= 1.7893e-002

Rho= 0.9997953

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	0.013	0.013	-20	0.000	-20					-
0.2	0.200	0.206	0.006	782	0.0 %	782					3
2.5	2.500	2.450	-0.050	10107	0.0 %	10107					-2
5.0	5.000	5.108	0.108	21149	0.0 %	21149					2
7.5	7.500	7.371	-0.129	30556	0.0 %	30556					-2
10.0	10.000	10.052	0.052	41693	0.0 %	41693					1

LB119462 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 31 Mar 2022 15:12:23					
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-20	-	0.00007471B	PPB		31 Mar 2022 15:14:08	S	Std
0.2 - 1	50.2	782	-	0.20007471B	PPB		31 Mar 2022 15:16:25	S	Std
2.5 - 1	52.5	10107	-	2.50007471B	PPB		31 Mar 2022 15:18:42	S	Std
5.0 - 1	55	21149	-	5.00007471B	PPB		31 Mar 2022 15:20:59	S	Std
7.5 - 1	57.5	30556	-	7.50007471B	PPB		31 Mar 2022 15:23:17	S	Std
10.0 - 1	511	41693	-	10.00007471B	PPB		31 Mar 2022 15:25:36	S	Std
ICV56 - 1	ICV56	16081	3.8879	-7471B	PPB		31 Mar 2022 15:28:35	U	SMPL
ICB56 - 1	ICB56	-18	0.0136	-7471B	PPB		31 Mar 2022 15:39:51	U	SMPL
CCV66 - 1	CCV66	21582	5.2118	-7471B	PPB		31 Mar 2022 15:42:06	U	SMPL
CCB66 - 1	CCB66	99	0.0417	-7471B	PPB		31 Mar 2022 15:47:03	U	SMPL
CRA - 1	CRA	750	0.1984	-7471B	PPB		31 Mar 2022 15:49:17	U	SMPL
HighStd - 1	HighStd	42187	10.1705	-7471B	PPB		31 Mar 2022 15:51:32	U	SMPL
ChkStd - 1	ChkStd	28780	6.9440	-7471B	PPB		31 Mar 2022 15:53:47	U	SMPL
PB143784BL - 1	PBS	-51	0.0056	-7471B	PPB		31 Mar 2022 15:58:46	U	SMPL
PB143784BS - 1	LCSS	15784	3.8164	-7471B	PPB		31 Mar 2022 16:01:01	U	SMPL
N1645-01 - 1	LOD-MDL-SOIL-01-QT1-2022	154	0.0550	-7471B	PPB		31 Mar 2022 16:10:55	U	SMPL
N1645-01RE - 1	LOD-MDL-SOIL-01-QT1-2022RE	573	0.1558	-7471B	PPB		31 Mar 2022 16:13:11	U	SMPL
N1645-03 - 1	MDL-SOIL-03-QT1-2022	181	0.0615	-7471B	PPB		31 Mar 2022 16:15:28	U	SMPL
N2195-01 - 1	TP-1-1-5FT	38842	9.3655	-7471B	PPB		31 Mar 2022 16:17:45	U	SMPL
N2195-03 - 1	TP-2-1-5FT	21745	5.2510	-7471B	PPB		31 Mar 2022 16:20:00	U	SMPL
CCV67 - 1	CCV67	20897	5.0469	-7471B	PPB		31 Mar 2022 16:22:23	U	SMPL
CCB67 - 1	CCB67	-317	-0.0584	-7471B	PPB		31 Mar 2022 16:24:42	U	SMPL
N2195-05 - 1	TP-3-1-5FT	20574	4.9692	-7471B	PPB		31 Mar 2022 16:27:01	U	SMPL
N2198-01 - 1	RO-VNJ-202	2033	0.5071	-7471B	PPB		31 Mar 2022 16:29:16	U	SMPL
N2199-01 - 1	RO-VNJ-209	2103	0.5240	-7471B	PPB		31 Mar 2022 16:31:35	U	SMPL
N2199-03 - 1	RO-ETGI-371	1312	0.3336	-7471B	PPB		31 Mar 2022 16:33:52	U	SMPL
N2202-01 - 1	MHE	399	0.1139	-7471B	PPB		31 Mar 2022 16:36:08	U	SMPL
N2202-04 - 1	MHF	740	0.1960	-7471B	PPB		31 Mar 2022 16:38:25	U	SMPL
N2202-04DUP - 1	MHFDUP	668	0.1787	-7471B	PPB		31 Mar 2022 16:40:41	U	SMPL
N2202-04MS - 1	MHFMS	17545	4.2402	-7471B	PPB		31 Mar 2022 16:45:40	U	SMPL
N2202-04MSD - 1	MHFMSD	19156	4.6279	-7471B	PPB		31 Mar 2022 16:50:30	U	SMPL
CCV68 - 1	CCV68	20661	4.9901	-7471B	PPB		31 Mar 2022 16:58:11	U	SMPL
CCB68 - 1	CCB68	-309	-0.0565	-7471B	PPB		31 Mar 2022 17:00:26	U	SMPL
N2202-04LX5 - 1		215	0.0696	-7471B	PPB		31 Mar 2022 17:02:45	U	SMPL
N2202-04A - 1		18830	4.5495	-7471B	PPB		31 Mar 2022 17:05:00	U	SMPL
CCV69 - 1	CCV69	21904	5.2893	-7471B	PPB		31 Mar 2022 17:07:16	U	SMPL
CCB69 - 1	CCB69	11	0.0205	-7471B	PPB		31 Mar 2022 17:09:34	U	SMPL