



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB119465

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP96131
Chlorine Calibration std, 0.1ppm	WP96126
Chlorine Calibration std, 0.2ppm	WP96127
Chlorine Calibration std, 0.8ppm	WP96128
Chlorine Calibration std, 0.0ppm	WP96125
Chlorine Calibration std, 1.6ppm	WP96129
Residual Chlorine Calibration and CCV std, 0	WP96130
Total Chlorine Powder Pillows	W2759

Intercept: 0.0080

Slope: 1.0071

Regression: 0.999948

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		03/31/2022	16:10
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	-1	03/31/2022	16:13
3	CAL3	0.2	1	0.000	0.220	0.220	0.21	-5.5	03/31/2022	16:16
4	CAL4	0.4	1	0.000	0.410	0.410	0.40	0.3	03/31/2022	16:19
5	CAL5	0.8	1	0.000	0.810	0.810	0.80	0.5	03/31/2022	16:22
6	CAL6	1.6	1	0.000	1.620	1.620	1.60	-0.1	03/31/2022	16:25



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB119465

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.3990	03/31/2022	16:28
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	03/31/2022	16:31
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.4090	03/31/2022	16:34
4	CCB1				1	0.0000	0.0100	0.0100	0.0020	03/31/2022	16:37
5	LB119465BL	50	50		1	0.0000	0.0100	0.0100	0.0020	03/31/2022	16:40
6	LB119465BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.3990	03/31/2022	16:43
7	N2221-01	50	50		1	0.0000	0.8500	0.8500	0.8360	03/31/2022	16:46
8	N2221-01DUP	50	50		1	0.0000	0.8600	0.8600	0.8460	03/31/2022	16:49
9	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3890	03/31/2022	16:52
10	CCB2				1	0.0000	0.0000	0.0000	-0.0080	03/31/2022	16:55

WORKLIST(Hardcopy Internal Chain)

LB 118465

WorkList Name : RESCHL-033122 WorkList ID : 158266 Department : Wet-Chemistry Date : 03-31-2022 16:14:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
3/30/2022	Water	N2221-01	Residual Chlorine	Cool 4 deg C	STER08	O11	POOL-WATER	03/30/2022	SM4500 Cl