



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119466

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	03/31/2022	17:15
2	CAL2	1	Water	NA	NA	20.6	7.00	03/31/2022	17:16
3	CAL3	1	Water	NA	NA	20.6	10.03	03/31/2022	17:19
4	ICV	1	Water	NA	NA	20.7	6.98	03/31/2022	17:20
5	CCV1	1	Water	NA	NA	20.7	2.01	03/31/2022	17:22
6	N2213-01	1	Solid	20.03	20	21.1	6.80	03/31/2022	17:25
7	N2214-01	1	Solid	20.01	20	20.9	5.51	03/31/2022	17:26
8	N2219-02	1	Solid	20.01	20	20.6	8.20	03/31/2022	17:30
9	N2219-02DUP	1	Solid	20.02	20	20.7	8.21	03/31/2022	17:31
10	CCV2	1	Water	NA	NA	20.7	12.03	03/31/2022	17:33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2219 WorkList ID : 158267 Department : Wet-Chemistry Date : 03-31-2022 17:06:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/31/2022	Solid	N2213-01	Corrosivity	Cool 4 deg C	PSEG03	O11	SOIL-DRUM	03/31/2022	9045D
03/31/2022	Solid	N2214-01	Corrosivity	Cool 4 deg C	PSEG03	O11	304	03/31/2022	9045D
03/31/2022	Solid	N2219-02	Corrosivity	Cool 4 deg C	PSEG03	O11	TP-3	03/31/2022	9045D

Date/Time 03/31/22 17:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 03/31/22 18:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]