

LB119481


245.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

31423



A= 0.0000e+000

B= 3.1271e-004

C= 1.0789e-003

Rho= 0.9995764

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.003	-0.003	-12	0.000	-12					-16
0.05	0.050	0.042	-0.008	130	0.0 %	130	0				-5
0.2	0.200	0.191	-0.009	606	0.0 %	606					-2
2.5	2.500	2.447	-0.053	7822	0.0 %	7822					1
5.0	5.000	5.026	0.026	16069	0.0 %	16069					3
7.5	7.500	7.720	0.220	24683	0.0 %	24683					-2
10.0	10.000	9.827	-0.173	31423	0.0 %	31423					

LB119481 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Apr 2022 14:14:13

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	245.1	PPB	01 Apr 2022 14:18:05	S	Std
0.05 - 1	50.65	130	-	0.0500	245.1	PPB	01 Apr 2022 14:22:57	S	Std
0.2 - 1	50.2	606	-	0.2000	245.1	PPB	01 Apr 2022 14:25:13	S	Std
2.5 - 1	52.3	7822	-	2.5000	245.1	PPB	01 Apr 2022 14:27:31	S	Std
5.0 - 1	55	16069	-	5.0000	245.1	PPB	01 Apr 2022 14:29:47	S	Std
7.5 - 1	57.5	24683	-	7.5000	245.1	PPB	01 Apr 2022 14:32:03	S	Std
10.0 - 1	57.6	31423	-	10.0000	245.1	PPB	01 Apr 2022 14:39:33	S	Std
ICV57 - 1	ICV57	12733	3.9828	-245.1		PPB	01 Apr 2022 14:43:55	U	SMPL
ICB57 - 1	ICB57	-75	-0.0224	-245.1		PPB	01 Apr 2022 14:48:43	U	SMPL
CCV70 - 1	CCV70	16342	5.1114	-245.1		PPB	01 Apr 2022 14:50:58	U	SMPL
CCB70 - 1	CCB70	-65	-0.0192	-245.1		PPB	01 Apr 2022 14:55:47	U	SMPL
CRA - 1	CRA	160	0.0511	-245.1		PPB	01 Apr 2022 15:00:33	U	SMPL
HighStd - 1	HighStd	31388	9.8165	-245.1		PPB	01 Apr 2022 15:02:47	U	SMPL
ChkStd - 1	ChkStd	23869	7.4652	-245.1		PPB	01 Apr 2022 15:05:03	U	SMPL
PB143811BL - 1	PBW	-40	-0.0114	-245.1		PPB	01 Apr 2022 15:09:57	U	SMPL
PB143811BS - 1	LCSW	12494	3.9081	-245.1		PPB	01 Apr 2022 15:12:12	U	SMPL
N1645-10 - 1	LOD-MDL-WATER-04-QT1-2022	134	0.0430	-245.1		PPB	01 Apr 2022 15:17:12	U	SMPL
N1645-10RE - 1	LOD-MDL-WATER-04-QT1-2022RE	495	0.1559	-245.1		PPB	01 Apr 2022 15:19:28	U	SMPL
N1645-12 - 1	MDL-WATER-06-QT1-2022	165	0.0527	-245.1		PPB	01 Apr 2022 15:21:45	U	SMPL
CCV71 - 1	CCV71	16772	5.2459	-245.1		PPB	01 Apr 2022 15:24:02	U	SMPL
CCB71 - 1	CCB71	-38	-0.0108	-245.1		PPB	01 Apr 2022 15:28:46	U	SMPL