

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 04/02/2022

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 04/01/2022 17:25

DATE & TIME OUT1: 04/02/2022 08:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 103 °C

DATE & TIME IN2: 04/02/2022 08:30

DATE & TIME OUT2: 04/02/2022 10:00

Run Number: LB119493

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB119493BL	LB119493BL	98.6371	98.6371	100	98.6371	98.6371	98.6371	0.0000	0
2	N2229-03	AQUEOUS-DRUM	85.0738	85.0738	100	85.3190	85.3190	85.3190	0.2452	2452
3	N2229-03DUP	AQUEOUS-DRUMDUP	63.8126	63.8126	100	64.0630	64.0630	64.0630	0.2504	2504

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WB 114493

WorkList Name : TS N2229 WorkList ID : 158308 Department : Wet-Chemistry Date : 04-01-2022 17:05:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/08/2022	Water	N2229-03	TS	Cool 4 deg C	PSEG03	O11	AQUEOUS-DRUM	04/01/2022	SM2540 B

Date/Time 04/01/2022 17:10
 Raw Sample Received by: CP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 04/01/2022 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)