



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 04/01/2022 17:25 TEMP1 OUT: 103 °C 04/02/2022 08:00
TEMP2 IN: 103 °C 04/02/2022 08:30 TEMP2 OUT: 103 °C 04/02/2022 10:00
TEMP3 IN: 540 °C 04/02/2022 10:30 TEMP3 OUT: 540 °C 04/02/2022 12:00
TEMP4 IN: 560 °C 04/02/2022 12:30 TEMP4 OUT: 560 °C 04/02/2022 14:00

Run Number: LB119494

SUPERVISOR: apatel

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB119494BL	98.6371	98.6371	100	98.6371	98.6371	98.6371	98.6371	0.0000	0
2	N2229-03	85.0738	85.0738	100	85.3190	85.319	85.2947	85.2947	0.0243	243
3	N2229-03DUP	63.8126	63.8126	100	64.0630	64.063	64.0378	64.0378	0.0252	252

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB119494

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS N2229 WorkList ID : 158309 Department : Wet-Chemistry Date : 04-01-2022 17:07:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/08/2022	Water	N2229-03	TVS	Cool 4 deg C	PSEG03	O11	AQUEOUS-DRUM	04/01/2022	SM2540 G

Date/Time 04/01/2022 17:10
Raw Sample Received by: JPCWC1
Raw Sample Relinquished by: CP SM

Date/Time 04/01/2022 18:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JPCWC1