



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119496

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/02/2022	09:15
2	CAL2	1	Water	NA	NA	20.1	7.00	04/02/2022	09:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/02/2022	09:18
4	ICV	1	Water	NA	NA	20.2	6.99	04/02/2022	09:20
5	CCV1	1	Water	NA	NA	20.2	2.01	04/02/2022	09:22
6	N2218-01	1	Water	NA	NA	20.2	7.05	04/02/2022	09:25
7	N2229-03	1	Water	NA	NA	20.1	6.89	04/02/2022	09:26
8	N2229-03DUP	1	Water	NA	NA	20.1	6.90	04/02/2022	09:27
9	CCV2	1	Water	NA	NA	20.3	12.03	04/02/2022	09:30

VB 119496

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2246 WorkList ID : 158307 Department : Wet-Chemistry Date : 04-02-2022 09:09:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/31/2022	Water	N2218-01	pH	Cool 4 deg C	PSEG03	O11	TOTE	03/31/2022	9040C
04/01/2022	Water	N2229-03	pH	Cool 4 deg C	PSEG03	O11	AQUEOUS-DRUM	04/01/2022	9040C

Date/Time 04/02/22 09:10
Raw Sample Received by: JD (CWC)
Raw Sample Relinquished by: CP SM

Date/Time 04/02/22 12:20
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JD (CWC)