


LB119502

7470A

INSTRUMENT ID: CV1

 μ Abs.:

35999



A= 0.0000e+000

B= 2.7603e-004

C= 4.3699e-002

Rho= 0.9999443

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/1)
0.0	0.000	0.039	0.039	-16	0.000	-16					-
0.2	0.200	0.202	0.002	574	0.0 %	574					1
2.5	2.500	2.437	-0.063	8671	0.0 %	8671					-3
5.0	5.000	4.986	-0.014	17904	0.0 %	17904					0
7.5	7.500	7.555	0.055	27212	0.0 %	27212					1
10.0	10.000	9.981	-0.019	35999	0.0 %	35999					0

LB119502 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 04 Apr 2022 08:54:22					
Sample ID	Extended ID	μ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-16	-	0.000	7470A	PPB	04 Apr 2022 09:18:24	S	Std
0.2 - 1	50.2	574	-	0.200	7470A	PPB	04 Apr 2022 09:20:41	S	Std
2.5 - 1	52.5	8671	-	2.500	7470A	PPB	04 Apr 2022 09:22:58	S	Std
5.0 - 1	55	17904	-	5.000	7470A	PPB	04 Apr 2022 09:25:15	S	Std
7.5 - 1	57.5	27212	-	7.500	7470A	PPB	04 Apr 2022 09:30:09	S	Std
10.0 - 1	57.8	35999	-	10.000	7470A	PPB	04 Apr 2022 09:32:24	S	Std
ICV58 - 1	ICV58	13344	3.7271	-	7470A	PPB	04 Apr 2022 09:35:08	U	SMPL
ICB58 - 1	ICB58	-142	0.0045	-	7470A	PPB	04 Apr 2022 09:39:58	U	SMPL
CCV72 - 1	CCV72	17807	4.9590	-	7470A	PPB	04 Apr 2022 09:42:13	U	SMPL
CCB72 - 1	CCB72	-73	0.0235	-	7470A	PPB	04 Apr 2022 09:47:06	U	SMPL
CRA - 1	CRA	539	0.1925	-	7470A	PPB	04 Apr 2022 09:49:20	U	SMPL
HighStd - 1	HighStd	35900	9.9532	-	7470A	PPB	04 Apr 2022 09:51:35	U	SMPL
ChkStd - 1	ChkStd	26091	7.2456	-	7470A	PPB	04 Apr 2022 09:53:51	U	SMPL
PB143780BL - 1	PBW	-299	-0.0388	-	7470A	PPB	04 Apr 2022 10:00:32	U	SMPL
PB143780BS - 1	LCSW	13725	3.8322	-	7470A	PPB	04 Apr 2022 10:02:49	U	SMPL
N2167-02 - 1	RAIL-PILE-1	-322	-0.0452	-	7470A	PPB	04 Apr 2022 10:05:05	U	SMPL
N2167-05 - 1	RAIL-PILE-2	-92	0.0183	-	7470A	PPB	04 Apr 2022 10:07:22	U	SMPL
N2202-02 - 1	MHE	-57	0.0280	-	7470A	PPB	04 Apr 2022 10:09:39	U	SMPL
N2202-05 - 1	MHF	-26	0.0365	-	7470A	PPB	04 Apr 2022 10:11:56	U	SMPL
N2202-05DUP - 1	MHFDUP	-37	0.0335	-	7470A	PPB	04 Apr 2022 10:14:13	U	SMPL
CCV73 - 1	CCV73	18007	5.0142	-	7470A	PPB	04 Apr 2022 10:16:29	U	SMPL
CCB73 - 1	CCB73	-335	-0.0488	-	7470A	PPB	04 Apr 2022 10:18:44	U	SMPL
N2202-05MS - 1	MHFMS	12667	3.5402	-	7470A	PPB	04 Apr 2022 10:21:01	U	SMPL
N2202-05MSD - 1	MHFMSD	13843	3.8648	-	7470A	PPB	04 Apr 2022 10:23:17	U	SMPL
PB143807BL - 1	PBW	-89	0.0191	-	7470A	PPB	04 Apr 2022 10:28:16	U	SMPL
PB143807BS - 1	LCSW	12667	3.5402	-	7470A	PPB	04 Apr 2022 10:30:31	U	SMPL
N2214-01 - 1	304	-288	-0.0358	-	7470A	PPB	04 Apr 2022 10:32:48	U	SMPL
N2215-07 - 1	COMP-1	-81	0.0213	-	7470A	PPB	04 Apr 2022 10:35:05	U	SMPL
N2215-08 - 1	COMP-2	-44	0.0316	-	7470A	PPB	04 Apr 2022 10:37:21	U	SMPL
N2219-02 - 1	TP-3	-20	0.0382	-	7470A	PPB	04 Apr 2022 10:39:38	U	SMPL
N2219-02DUP - 1	TP-3DUP	-16	0.0393	-	7470A	PPB	04 Apr 2022 10:41:55	U	SMPL
N2219-02MS - 1	TP-3MS	12123	3.3900	-	7470A	PPB	04 Apr 2022 10:44:10	U	SMPL
CCV74 - 1	CCV74	18121	5.0457	-	7470A	PPB	04 Apr 2022 10:46:26	U	SMPL
CCB74 - 1	CCB74	-387	-0.0631	-	7470A	PPB	04 Apr 2022 10:48:43	U	SMPL
N2219-02MSD - 1	TP-3MSD	12176	3.4047	-	7470A	PPB	04 Apr 2022 10:50:59	U	SMPL
PB143809BL - 1	PBW	-279	-0.0333	-	7470A	PPB	04 Apr 2022 10:53:15	U	SMPL
PB143809BS - 1	LCSW	14575	4.0669	-	7470A	PPB	04 Apr 2022 10:58:10	U	SMPL
N2215-09 - 1	GW-1	-350	-0.0529	-	7470A	PPB	04 Apr 2022 11:00:25	U	SMPL
N2215-09DUP - 1	GW-1DUP	-117	0.0114	-	7470A	PPB	04 Apr 2022 11:02:42	U	SMPL
N2215-09MS - 1	GW-1MS	12918	3.6095	-	7470A	PPB	04 Apr 2022 11:04:58	U	SMPL
N2215-09MSD - 1	GW-1MSD	11446	3.2032	-	7470A	PPB	04 Apr 2022 11:07:14	U	SMPL
N2215-10 - 1	GW-2	35	0.0534	-	7470A	PPB	04 Apr 2022 11:09:33	U	SMPL
N2215-11 - 1	GW-3	-62	0.0266	-	7470A	PPB	04 Apr 2022 11:11:51	U	SMPL
PB143837BL - 1	PBW	-47	0.0307	-	7470A	PPB	04 Apr 2022 11:14:06	U	SMPL
CCV75 - 1	CCV75	18687	5.2019	-	7470A	PPB	04 Apr 2022 11:16:22	U	SMPL
CCB75 - 1	CCB75	-370	-0.0584	-	7470A	PPB	04 Apr 2022 11:18:37	U	SMPL
PB143837BS - 1	LCSW	14607	4.0757	-	7470A	PPB	04 Apr 2022 11:20:55	U	SMPL
N2197-01 - 1	DEWATERING-EFFLUENT-033022	-309	-0.0416	-	7470A	PPB	04 Apr 2022 11:23:10	U	SMPL
N2215-03 - 1	GW-1	1407	0.4321	-	7470A	PPB	04 Apr 2022 11:30:25	U	SMPL
N2215-03DUP - 1	GW-1DUP	1354	0.4174	-	7470A	PPB	04 Apr 2022 11:45:44	U	SMPL
N2215-03MS - 1	GW-1MS	13979	3.9023	-	7470A	PPB	04 Apr 2022 11:47:59	U	SMPL
N2215-03MSD - 1	GW-1MSD	14597	4.0729	-	7470A	PPB	04 Apr 2022 11:50:15	U	SMPL
N2215-04 - 1	GW-2	2179	0.6452	-	7470A	PPB	04 Apr 2022 11:52:32	U	SMPL
N2215-05 - 1	GW-3	1029	0.3277	-	7470A	PPB	04 Apr 2022 11:54:50	U	SMPL
N2218-01 - 1	TOTE	1426	0.4373	-	7470A	PPB	04 Apr 2022 11:57:07	U	SMPL
N2220-18 - 1	FB0331	-29	0.0357	-	7470A	PPB	04 Apr 2022 11:59:23	U	SMPL
CCV76 - 1	CCV76	18827	5.2405	-	7470A	PPB	04 Apr 2022 12:01:38	U	SMPL
CCB76 - 1	CCB76	-141	0.0048	-	7470A	PPB	04 Apr 2022 12:03:54	U	SMPL
PB143752TB - 1	PB143752TB	-29	0.0357	-	7470A	PPB	04 Apr 2022 12:06:11	U	SMPL
PB143785TB - 1	PB143785TB	-3	0.0429	-	7470A	PPB	04 Apr 2022 12:08:27	U	SMPL
PB143787TB - 1	PB143787TB	1	0.0440	-	7470A	PPB	04 Apr 2022 12:10:43	U	SMPL
N2202-05LX5 - 1		-5	0.0423	-	7470A	PPB	04 Apr 2022 12:13:00	U	SMPL
N2202-05A - 1		13192	3.6851	-	7470A	PPB	04 Apr 2022 12:15:16	U	SMPL
N2219-02LX5 - 1		-94	0.0178	-	7470A	PPB	04 Apr 2022 12:17:32	U	SMPL
N2219-02A - 1		14460	4.0351	-	7470A	PPB	04 Apr 2022 12:19:48	U	SMPL
N2215-09LX5 - 1		-147	0.0031	-	7470A	PPB	04 Apr 2022 12:22:05	U	SMPL
N2215-09A - 1		15071	4.2038	-	7470A	PPB	04 Apr 2022 12:24:23	U	SMPL
CCV77 - 1	CCV77	18435	5.1323	-	7470A	PPB	04 Apr 2022 12:26:39	U	SMPL
CCB77 - 1	CCB77	-250	-0.0253	-	7470A	PPB	04 Apr 2022 12:28:57	U	SMPL
N2215-03LX5 - 1		231	0.1075	-	7470A	PPB	04 Apr 2022 12:31:15	U	SMPL
N2215-03A - 1		16332	4.5518	-	7470A	PPB	04 Apr 2022 12:41:22	U	SMPL
CCV78 - 1	CCV78	18640	5.1889	-	7470A	PPB	04 Apr 2022 12:43:38	U	SMPL
CCB78 - 1	CCB78	-307	-0.0410	-	7470A	PPB	04 Apr 2022 12:45:55	U	SMPL