



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/5/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 04/04/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 04/05/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119509

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2250-01	040122TB01-PCB	1	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N2250-02	040122TB02-PCB	2	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N2250-03	040122TB03-PCB	3	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N2251-01	20220353-MANASQUAN-TUBE -9-COMPOSITE	4	1.14	8.80	9.94	5.61	50.8	SLUDGE SAMPLE
N2251-02	20220354-MANASQUAN-TUBE -15-COMPOSITE	5	1.18	8.78	9.96	4.76	40.8	SLUDGE SAMPLE
N2251-03	20220355-MANASQUAN-TUBE -19-COMPOSITE	6	1.16	8.73	9.89	4.27	35.6	SLUDGE SAMPLE
N2251-04	20220356-MANASQUAN-TUBE -18-COMPOSITE	7	1.17	8.81	9.98	4.03	32.5	SLUDGE SAMPLE
N2252-01	OK-01-040422	25	1.13	8.54	9.67	8.61	87.6	
N2252-02	OK-01-040422-EPH-2	26	1.13	8.54	9.67	8.61	87.6	
N2253-01	101221	8	1.00	1.00	2.00	2.00	100.0	RAGS
N2253-02	31292	9	1.00	1.00	2.00	2.00	100.0	RAGS
N2253-03	31396	10	1.00	1.00	2.00	2.00	100.0	RAGS
N2255-01	31496	11	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2255-02	31495	12	1.14	8.77	9.91	9.84	99.2	
N2256-01	SU-02-040422	27	1.12	8.46	9.58	9.01	93.3	
N2256-02	SU-02-040422-EPH-2	28	1.11	8.78	9.89	9.33	93.6	
N2256-03	SU-03-040422	29	1.13	8.53	9.66	8.54	86.9	
N2256-04	SU-03-040422-EPH-2	30	1.14	8.83	9.97	9.02	89.2	
N2257-01	TP-1	13	1.15	8.80	9.95	8.91	88.2	
N2257-02	TP-1-EPH-2	14	1.18	8.48	9.66	8.34	84.4	
N2257-03	TP-2	15	1.19	8.58	9.77	8.73	87.9	
N2257-04	TP-2-EPH-2	16	1.18	8.59	9.77	8.81	88.8	
N2257-05	TP-3	17	1.16	8.75	9.91	9.00	89.6	
N2257-06	TP-3-EPH-2	18	1.11	8.72	9.83	8.96	90.0	
N2257-07	TP-4	19	1.14	8.66	9.8	9.06	91.5	
N2257-08	TP-4-EPH-2	20	1.15	8.75	9.9	9.37	93.9	



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/5/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 04/04/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 04/05/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119509

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2257-09	TP-5	21	1.18	8.54	9.72	8.53	86.1	
N2257-10	TP-5-EPH-2	22	1.16	8.48	9.64	8.46	86.1	
N2257-11	TP-6	23	1.12	8.47	9.59	8.25	84.2	
N2257-12	TP-6-EPH-2	24	1.18	8.72	9.9	8.73	86.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WB19509

WORKLIST(Hardcopy Internal Chain)

: %1-040422

WorkList ID : 158338

Department : Wet-Chemistry

Date : 04-04-2022

Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date
					Storage Location	Customer Sample	
Solid	N2250-01	Percent Solids	Cool 4 deg C	CLOU03	N21	040122TB01-PCB	04/01/2022
Solid	N2250-02	Percent Solids	Cool 4 deg C	CLOU03	N21	040122TB02-PCB	04/01/2022
Solid	N2250-03	Percent Solids	Cool 4 deg C	CLOU03	N21	040122TB03-PCB	04/01/2022
Solid	N2251-01	Percent Solids	Cool 4 deg C	AQUA04	O11	20220353-MANASQUAN-TUBE-9-COMPOSITE	03/29/2022
Solid	N2251-02	Percent Solids	Cool 4 deg C	AQUA04	O11	20220354-MANASQUAN-TUBE-15-COMPOSITE	03/29/2022
Solid	N2251-03	Percent Solids	Cool 4 deg C	AQUA04	O11	20220355-MANASQUAN-TUBE-19-COMPOSITE	03/29/2022
Solid	N2251-04	Percent Solids	Cool 4 deg C	AQUA04	O11	20220356-MANASQUAN-TUBE-18-COMPOSITE	03/29/2022
Solid	N2253-01	Percent Solids	Cool 4 deg C	PSEG03	O11	101221	04/04/2022
Solid	N2253-02	Percent Solids	Cool 4 deg C	PSEG03	O11	31292	04/04/2022
Solid	N2253-03	Percent Solids	Cool 4 deg C	PSEG03	O11	31396	04/04/2022
Solid	N2255-01	Percent Solids	Cool 4 deg C	PSEG03	O11	31496	04/04/2022
Solid	N2255-02	Percent Solids	Cool 4 deg C	PSEG03	O11	31495	04/04/2022
Solid	N2257-01	Percent Solids	Cool 4 deg C	PSEG03	O11	TP-1	04/04/2022
Solid	N2257-08	Percent Solids	Cool 4 deg C	PSEG03	O11	TP-4-EPH-2	04/04/2022

70
1054-878
13

WORKLIST(Hardcopy Internal Chain)

WorkList ID : 158338 Department : Wet-Chemistry Date : 04-04-2022

Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date
					Storage Location	Customer Sample	
Solid	N2257-05	Percent Solids	Cool 4 deg C	PSEG03	O11	TP-3	04/04/2022 C
Solid	N2257-06	Percent Solids	Cool 4 deg C	PSEG03	O11	TP-3-EPH-2	04/04/2022 C
Solid	N2257-07	Percent Solids	Cool 4 deg C	PSEG03	O11	TP-4	04/04/2022 C
Solid	N2252-01	Percent Solids	Cool 4 deg C	PSEG05	O11	OK-01-040422	04/04/2022 C
Solid	N2252-02	Percent Solids	Cool 4 deg C	PSEG05	O11	OK-01-040422-EPH-2	04/04/2022 C
Solid	N2256-01	Percent Solids	Cool 4 deg C	PSEG05	O11	SU-02-040422	04/04/2022 C
Solid	N2256-02	Percent Solids	Cool 4 deg C	PSEG05	O11	SU-02-040422-EPH-2	04/04/2022 C
Solid	N2256-03	Percent Solids	Cool 4 deg C	PSEG05	O11	SU-03-040422	04/04/2022 C
Solid	N2256-04	Percent Solids	Cool 4 deg C	PSEG05	O11	SU-03-040422-EPH-2	04/04/2022 C