



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119516

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/04/2022	17:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/04/2022	17:16
3	CAL3	1	Water	NA	NA	20.3	10.03	04/04/2022	17:18
4	ICV	1	Water	NA	NA	20.4	6.99	04/04/2022	17:20
5	CCV1	1	Water	NA	NA	20.3	2.01	04/04/2022	17:22
6	N2253-05	1	Water	NA	NA	20.4	7.55	04/04/2022	17:34
7	N2253-05DUP	1	Water	NA	NA	20.5	7.56	04/04/2022	17:35
8	N2253-06	1	Water	NA	NA	20.6	6.74	04/04/2022	17:37
9	N2255-03	1	Water	NA	NA	20.6	6.45	04/04/2022	17:39
10	CCV2	1	Water	NA	NA	20.1	12.02	04/04/2022	17:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2253 WorkList ID : 158370 Department : Wet-Chemistry Date : 04-04-2022 17:05:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/04/2022	Water	N2253-05	pH	Cool 4 deg C	PSEG03	O11	41044	04/04/2022	9040C
04/04/2022	Water	N2253-06	pH	Cool 4 deg C	PSEG03	O11	2822S	04/04/2022	9040C
04/04/2022	Water	N2255-03	pH	Cool 4 deg C	PSEG03	O11	31496	04/04/2022	9040C

Date/Time 04/04/22 17:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/04/22 18:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]