

LB119522


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35348



A= 0.0000e+000

B= 2.8178e-004

C= 4.5146e-002

Rho= 0.9995368

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o D
0.0	0.000	0.039	0.039	-22	0.000	-22					-
0.2	0.200	0.204	0.004	562	0.0 %	562					2
2.5	2.500	2.572	0.072	8968	0.0 %	8968					3
5.0	5.000	4.764	-0.236	16748	0.0 %	16748					-5
7.5	7.500	7.616	0.116	26867	0.0 %	26867					2
10.0	10.000	10.005	0.005	35348	0.0 %	35348					0

LB119522 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 05 Apr 2022 11:33:18				Date	Type	Type
Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units				
0.0 - 1	50	-22	-	0.0000	7470A	PPB		05 Apr 2022 11:44:18	S	Std
0.2 - 1	50.2	562	-	0.2000	7470A	PPB		05 Apr 2022 11:46:35	S	Std
2.5 - 1	52.5	8968	-	2.5000	7470A	PPB		05 Apr 2022 11:48:52	S	Std
5.0 - 1	55	16748	-	5.0000	7470A	PPB		05 Apr 2022 11:51:09	S	Std
7.5 - 1	57.5	26867	-	7.5000	7470A	PPB		05 Apr 2022 11:53:25	S	Std
10.0 - 1	57.8	35348	-	10.0000	7470A	PPB		05 Apr 2022 11:58:16	S	Std
ICV60 - 1	ICV60	14864	4.2335		7470A	PPB		05 Apr 2022 12:03:30	U	SMPL
ICB60 - 1	ICB60	-422	-0.0738		7470A	PPB		05 Apr 2022 12:05:45	U	SMPL
CCV85 - 1	CCV85	18350	5.2158		7470A	PPB		05 Apr 2022 12:08:02	U	SMPL
CCB85 - 1	CCB85	-388	-0.0642		7470A	PPB		05 Apr 2022 12:10:17	U	SMPL
CRA - 1	CRA	502	0.1866		7470A	PPB		05 Apr 2022 12:12:34	U	SMPL
HighStd - 1	HighStd	35955	10.1765		7470A	PPB		05 Apr 2022 12:14:49	U	SMPL
ChkStd - 1	ChkStd	23085	6.5500		7470A	PPB		05 Apr 2022 12:18:56	U	SMPL
PB143858BL - 1	PBW	-450	-0.0817		7470A	PPB		05 Apr 2022 12:21:11	U	SMPL
PB143858BS - 1	LCSW	13735	3.9154		7470A	PPB		05 Apr 2022 12:26:01	U	SMPL
N2207-09 - 1	SPLP-LEACHATE-AMENDED-096-21-01	-354	-0.0546		7470A	PPB		05 Apr 2022 12:28:17	U	SMPL
N2207-09DUP - 1	SPLP-LEACHATE-AMENDED-096-21-01DUP	17	0.0499		7470A	PPB		05 Apr 2022 12:30:35	U	SMPL
N2207-09MS - 1	SPLP-LEACHATE-AMENDED-096-21-01MS	11635	3.3236		7470A	PPB		05 Apr 2022 12:41:17	U	SMPL
N2207-09MSD - 1	SPLP-LEACHATE-AMENDED-096-21-01MSD	13022	3.7145		7470A	PPB		05 Apr 2022 12:43:33	U	SMPL
N2207-10 - 1	SPLP-LEACHATE-AMENDED-096-21-02	-138	0.0063		7470A	PPB		05 Apr 2022 12:45:49	U	SMPL
CCV86 - 1	CCV86	18806	5.3443		7470A	PPB		05 Apr 2022 12:48:41	U	SMPL
CCB86 - 1	CCB86	-400	-0.0676		7470A	PPB		05 Apr 2022 12:50:56	U	SMPL
N2207-11 - 1	SPLP-LEACHATE-AMENDED-096-21-03	10	0.0480		7470A	PPB		05 Apr 2022 12:53:13	U	SMPL
N2207-12 - 1	SPLP-LEACHATE-AMENDED-098-21-01	28	0.0530		7470A	PPB		05 Apr 2022 12:55:28	U	SMPL
PB143859BL - 1	PBW	26	0.0525		7470A	PPB		05 Apr 2022 12:57:44	U	SMPL
PB143859BS - 1	LCSW	12590	3.5927		7470A	PPB		05 Apr 2022 13:02:35	U	SMPL
N2227-02 - 1	124055	-224	-0.0180		7470A	PPB		05 Apr 2022 13:04:50	U	SMPL
N2227-02DUP - 1	124055DUP	-45	0.0325		7470A	PPB		05 Apr 2022 13:07:08	U	SMPL
N2227-02MS - 1	124055MS	12015	3.4307		7470A	PPB		05 Apr 2022 13:09:24	U	SMPL
N2227-02MSD - 1	124055MSD	13057	3.7243		7470A	PPB		05 Apr 2022 13:11:41	U	SMPL
N2227-04 - 1	124035	-99	0.0173		7470A	PPB		05 Apr 2022 13:14:00	U	SMPL
N2229-02 - 1	SOIL-DRUM	20	0.0508		7470A	PPB		05 Apr 2022 13:16:16	U	SMPL
CCV87 - 1	CCV87	17553	4.9912		7470A	PPB		05 Apr 2022 13:18:32	U	SMPL
CCB87 - 1	CCB87	-415	-0.0718		7470A	PPB		05 Apr 2022 13:20:47	U	SMPL
N2236-01 - 1	COMP-1	7	0.0471		7470A	PPB		05 Apr 2022 13:23:05	U	SMPL
N2236-02 - 1	COMP-2	64	0.0632		7470A	PPB		05 Apr 2022 13:25:20	U	SMPL
N2236-03 - 1	RO-VNJ-204	56	0.0609		7470A	PPB		05 Apr 2022 13:27:36	U	SMPL
N2239-02 - 1	TP-4	72	0.0654		7470A	PPB		05 Apr 2022 13:29:52	U	SMPL
N2246-09 - 1	WC-8-9-2-1	625	0.2213		7470A	PPB		05 Apr 2022 13:32:08	U	SMPL
N2246-10 - 1	WC-8-9-2-2	476	0.1793		7470A	PPB		05 Apr 2022 13:34:24	U	SMPL
N2246-11 - 1	WC-8-9-2-3	694	0.2407		7470A	PPB		05 Apr 2022 13:36:40	U	SMPL
N2246-12 - 1	WC-8-9-3-1	477	0.1796		7470A	PPB		05 Apr 2022 13:38:57	U	SMPL
PB143860BL - 1	PBW	59	0.0618		7470A	PPB		05 Apr 2022 13:41:14	U	SMPL
PB143860BS - 1	LCSW	11893	3.3963		7470A	PPB		05 Apr 2022 13:43:30	U	SMPL
CCV88 - 1	CCV88	17441	4.9596		7470A	PPB		05 Apr 2022 13:45:45	U	SMPL
CCB88 - 1	CCB88	-375	-0.0605		7470A	PPB		05 Apr 2022 13:48:02	U	SMPL
N2236-04 - 1	17140	104	0.0745		7470A	PPB		05 Apr 2022 13:50:19	U	SMPL
N2236-04DUP - 1	17140DUP	145	0.0860		7470A	PPB		05 Apr 2022 13:52:35	U	SMPL
N2236-04MS - 1	17140MS	11952	3.4129		7470A	PPB		05 Apr 2022 13:54:50	U	SMPL
N2236-04MSD - 1	17140MSD	12474	3.5600		7470A	PPB		05 Apr 2022 13:57:06	U	SMPL
N2236-05 - 1	14116	371	0.1497		7470A	PPB		05 Apr 2022 13:59:23	U	SMPL
PB143861BL - 1	PBW	-1	0.0449		7470A	PPB		05 Apr 2022 14:13:08	U	SMPL
PB143861BS - 1	LCSW	11729	3.3501		7470A	PPB		05 Apr 2022 14:15:24	U	SMPL
N2229-03 - 1	AQUEOUS-DRUM	1420	0.4453		7470A	PPB		05 Apr 2022 14:25:17	U	SMPL
N2229-03DUP - 1	AQUEOUS-DRUMDUP	1290	0.4086		7470A	PPB		05 Apr 2022 14:27:34	U	SMPL
N2229-03MS - 1	AQUEOUS-DRUMMS	15395	4.3831		7470A	PPB		05 Apr 2022 14:29:50	U	SMPL
CCV89 - 1	CCV89	17693	5.0306		7470A	PPB		05 Apr 2022 14:32:06	U	SMPL
CCB89 - 1	CCB89	-171	-0.0030		7470A	PPB		05 Apr 2022 14:34:22	U	SMPL
N2229-03MSD - 1	AQUEOUS-DRUMMSD	15816	4.5017		7470A	PPB		05 Apr 2022 14:36:40	U	SMPL
N2237-08 - 1	FB-0401	-128	0.0091		7470A	PPB		05 Apr 2022 14:38:55	U	SMPL
PB143858TB - 1	PB143818TB	-27	0.0375		7470A	PPB		05 Apr 2022 14:41:13	U	SMPL
PB143823TB - 1	PB143823TB	68	0.0643		7470A	PPB		05 Apr 2022 14:43:28	U	SMPL
PB143843TB - 1	PB143843TB	40	0.0564		7470A	PPB		05 Apr 2022 14:45:44	U	SMPL
CCV90 - 1	CCV90	17732	5.0416		7470A	PPB		05 Apr 2022 14:48:00	U	SMPL
CCB90 - 1	CCB90	-286	-0.0354		7470A	PPB		05 Apr 2022 14:50:17	U	SMPL
N2207-09LX5 - 1		7	0.0471		7470A	PPB		05 Apr 2022 14:52:35	U	SMPL
N2207-09A - 1		14574	4.1518		7470A	PPB		05 Apr 2022 14:54:52	U	SMPL
N2227-02LX5 - 1		-130	0.0085		7470A	PPB		05 Apr 2022 14:57:08	U	SMPL
N2227-02A - 1		14714	4.1912		7470A	PPB		05 Apr 2022 14:59:25	U	SMPL
N2236-04LX5 - 1		-83	0.0218		7470A	PPB		05 Apr 2022 15:01:41	U	SMPL
N2236-04A - 1		14709	4.1898		7470A	PPB		05 Apr 2022 15:03:58	U	SMPL
N2229-03LX5 - 1		640	0.2255		7470A	PPB		05 Apr 2022 15:06:14	U	SMPL
N2229-03A - 1		16608	4.7249		7470A	PPB		05 Apr 2022 15:08:32	U	SMPL
CCV91 - 1	CCV91	18174	5.1662		7470A	PPB		05 Apr 2022 15:13:31	U	SMPL
CCB91 - 1	CCB91	-364	-0.0574		7470A	PPB		05 Apr 2022 15:15:47	U	SMPL