



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB119529

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP95371
COD calibration std. 100 ppm	WP95373
COD calibration std. 10 ppm	WP95375
COD calibration std. 150 ppm	WP95372
COD calibration std. 50 ppm	WP95374
COD calibration std. 0 ppm	WP95376
COD Digestion Vials Low Level 0-150Mg/L	W2830
COD ICV-LCS std, 50ppm	WP96187
COD CCV std, 50ppm	WP96186

Temp In (C): 149	Date In: 04/05/2022	Time In: 10:20
Temp Out (C): 150	Date Out: 04/05/2022	Time Out: 12:20

Intercept: -0.4433

Slope: 1.0136

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.437		02/18/2022	13:20
2	CAL2	10	1	Water	9.000	9.317	6.8	02/18/2022	13:20
3	CAL3	50	1	Water	51.000	50.753	-1.5	02/18/2022	13:21
4	CAL4	100	1	Water	100.000	99.096	0.9	02/18/2022	13:21
5	CAL5	150	1	Water	152.000	150.398	-0.3	02/18/2022	13:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB119529

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	49.000	48.780	02/18/2022	13:22
2	ICB		NA	NA	1	Water	0.000	0.437	02/18/2022	13:23
3	CCV1	50	NA	NA	1	Water	51.000	50.753	04/05/2022	14:40
4	CCB1		NA	NA	1	Water	1.000	1.424	04/05/2022	14:40
5	LB119529BL		NA	NA	1	Water	1.000	1.424	04/05/2022	14:41
6	LB119529BS	50	NA	NA	1	Water	49.000	48.780	04/05/2022	14:41
7	N2246-15		NA	NA	1	Water	23.000	23.129	04/05/2022	14:42
8	N2246-15DUP		NA	NA	1	Water	24.000	24.115	04/05/2022	14:42
9	N2246-15MS	50	NA	NA	1	Water	73.000	72.458	04/05/2022	14:43
10	N2246-15MSD	50	NA	NA	1	Water	73.000	72.458	04/05/2022	14:43
11	N2246-16		NA	NA	1	Water	55.000	54.699	04/05/2022	14:44
12	N2246-17		NA	NA	1	Water	46.000	45.820	04/05/2022	14:44
13	N2246-18		NA	NA	1	Water	92.000	91.203	04/05/2022	14:45
14	CCV2	50	NA	NA	1	Water	51.000	50.753	04/05/2022	14:45
15	CCB2		NA	NA	1	Water	0.000	0.437	04/05/2022	14:46