



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119533

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/05/2022	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	04/05/2022	16:11
3	CAL3	1	Water	NA	NA	20.3	10.03	04/05/2022	16:12
4	ICV	1	Water	NA	NA	20.4	6.99	04/05/2022	16:15
5	CCV1	1	Water	NA	NA	20.4	2.01	04/05/2022	16:20
6	N2268-01	1	Water	NA	NA	20.1	5.17	04/05/2022	16:22
7	N2268-01DUP	1	Water	NA	NA	20.2	5.18	04/05/2022	16:23
8	N2269-01	1	Water	NA	NA	20.6	7.61	04/05/2022	16:25
9	N2269-02	1	Water	NA	NA	20.2	11.46	04/05/2022	16:30
10	N2269-03	1	Water	NA	NA	20.3	7.30	04/05/2022	16:33
11	CCV2	1	Water	NA	NA	20.4	12.03	04/05/2022	16:35

VB 119533

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2268

WorkList ID : 158406

Department : Wet-Chemistry

Date : 04-05-2022 15:52:45

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/23/2022	Water	N2268-01	pH	Cool 4 deg C	STER17	N12 JS9600	03/23/2022	9040C
04/05/2022	Water	N2269-01	pH	Cool 4 deg C	TULL02	O11 TP-6	04/05/2022	9040C
04/05/2022	Water	N2269-02	pH	Cool 4 deg C	TULL02	O11 TP-5	04/05/2022	9040C
04/05/2022	Water	N2269-03	pH	Cool 4 deg C	TULL02	O11 LP-2	04/05/2022	9040C

Date/Time

04/05/22 16:00

Raw Sample Received by:

JP (WC)

Raw Sample Relinquished by:

JP (SM)

Date/Time

04/05/22 17:20

Raw Sample Received by:

JP (SM)

Raw Sample Relinquished by:

JP (WC)