



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119540

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE (ICV)	W2843

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	04/05/2022	14:05
2	N2218-01		1	10	100.00	>212.0	>212.0	04/05/2022	14:18
3	N2229-03		1	9	100.00	>212.0	>212.0	04/05/2022	14:35
4	N2236-04		1	9	100.00	>212.0	>212.0	04/05/2022	14:49
5	N2236-05		1	10	100.00	>212.0	>212.0	04/05/2022	15:03
6	N2253-05		1	10	100.00	>212.0	>212.0	04/05/2022	15:20
7	N2253-06		1	9	100.00	>212.0	>212.0	04/05/2022	15:34
8	N2255-03		1	10	100.00	>212.0	>212.0	04/05/2022	15:50
9	N2269-01		1	10	100.00	>212.0	>212.0	04/05/2022	16:03
10	N2269-02		1	11	100.00	>212.0	>212.0	04/05/2022	16:18
11	N2269-03		1	10	100.00	>212.0	>212.0	04/05/2022	16:32
12	N2269-03DUP		1	10	100.00	>212.0	>212.0	04/05/2022	16:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 040522 WorkList ID : 158369 Department : Wet-Chemistry Date : 04-04-2022 16:36:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/08/2022	Water	N2269-01	Flash Point	Cool 4 deg C	TULL02	O11	TP-6	04/05/2022	1010B
04/08/2022	Water	N2269-02	Flash Point	Cool 4 deg C	TULL02	O11	TP-5	04/05/2022	1010B
04/08/2022	Water	N2269-03	Flash Point	Cool 4 deg C	TULL02	O11	LP-2	04/05/2022	1010B
04/07/2022	Water	N2218-01	Flash Point	Cool 4 deg C	PSEG03	O11	TOTE	03/31/2022	1010B
04/08/2022	Water	N2229-03	Flash Point	Cool 4 deg C	PSEG03	O11	AQUEOUS-DRUM	04/01/2022	1010B
04/08/2022	Water	N2236-04	Flash Point	Cool 4 deg C	PSEG03	O11	17140	04/01/2022	1010B
04/08/2022	Water	N2236-05	Flash Point	Cool 4 deg C	PSEG03	O11	14116	04/01/2022	1010B
04/07/2022	Water	N2253-05	Flash Point	Cool 4 deg C	PSEG03	O11	41044	04/04/2022	1010B
04/07/2022	Water	N2253-06	Flash Point	Cool 4 deg C	PSEG03	O11	2822S	04/04/2022	1010B
04/07/2022	Water	N2255-03	Flash Point	Cool 4 deg C	PSEG03	O11	31496	04/04/2022	1010B

Date/Time 4/5/22 13:10
 Raw Sample Received by: AL (LOC)
 Raw Sample Relinquished by: CR SM

Date/Time 4/5/22 16:15
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: AL (LOC)