



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119543

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.4	04/06/2022	08:45
2	N2253-04		1	9	100.00	>212.0	>212.0	04/06/2022	09:09
3	N2255-01		1	9	100.00	>212.0	>212.0	04/06/2022	09:25
4	N2255-01DUP		1	9	100.00	>212.0	>212.0	04/06/2022	09:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 040622 WorkList ID : 158404 Department : Wet-Chemistry Date : 04-05-2022 13:30:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/07/2022	Solid	N2253-04	Flash Point	Cool 4 deg C	PSEG03	O11	2822A	04/04/2022	1010B
04/07/2022	Solid	N2255-01	Flash Point	Cool 4 deg C	PSEG03	O11	31496	04/04/2022	1010B

Date/Time 4/5/22 13:30
 Raw Sample Received by: AL (yey)
 Raw Sample Relinquished by: CD SR

Date/Time 4/5/22 16:15
 Raw Sample Received by: CD SR
 Raw Sample Relinquished by: AL (wcy)