



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119545

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/06/2022	10:45
2	CAL2	1	Water	NA	NA	20.3	7.00	04/06/2022	10:46
3	CAL3	1	Water	NA	NA	20.2	10.02	04/06/2022	10:47
4	ICV	1	Water	NA	NA	20.4	6.99	04/06/2022	10:48
5	CCV1	1	Water	NA	NA	20.3	2.01	04/06/2022	10:50
6	N2279-01	1	Water	NA	NA	20.2	8.41	04/06/2022	11:00
7	N2279-01DUP	1	Water	NA	NA	20.3	8.42	04/06/2022	11:01
8	N2283-01	1	Water	NA	NA	20.6	4.67	04/06/2022	11:05
9	N2283-02	1	Water	NA	NA	20.1	4.75	04/06/2022	11:07
10	CCV2	1	Water	NA	NA	20.4	12.03	04/06/2022	11:10

WORKLIST(Hardcopy Internal Chain)

101545

WorkList Name : ph n2280 WorkList ID : 158431 Department : Wet-Chemistry Date : 04-06-2022 10:44:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/05/2022	Water	N2279-01	pH C	Cool 4 deg C	ENV127	N22	DEWATERING-EFF-040522	04/05/2022	9040C
04/04/2022	Water	N2283-01	pH A	Cool 4 deg C	STER16	N13	MAIN-CELL	04/04/2022	9040C
04/05/2022	Water	N2283-02	pH A	Cool 4 deg C	STER16	N13	EXCELL	04/05/2022	9040C

Date/Time 04/06/22 10:45
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: CA SM

Date/Time 04/06/22 12:00
 Raw Sample Received by: CA SM
 Raw Sample Relinquished by: JDC