

LB119548


7470A

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

37246



A= 0.0000e+000

B= 2.7141e-004

C= -4.3273e-002

Rho= 0.9997932

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.077  | 0.077  | 443   | 0.000      | 443   |       |       |       |       | 16 |
| 0.2    | 0.200  | 0.231  | 0.031  | 1009  | 0.0 %      | 1009  |       |       |       |       | -5 |
| 2.5    | 2.500  | 2.367  | -0.133 | 8880  | 0.0 %      | 8880  |       |       |       |       | 1  |
| 5.0    | 5.000  | 5.025  | 0.025  | 18675 | 0.0 %      | 18675 |       |       |       |       | -1 |
| 7.5    | 7.500  | 7.435  | -0.065 | 27553 | 0.0 %      | 27553 |       |       |       |       | -1 |
| 10.0   | 10.000 | 10.066 | 0.066  | 37246 | 0.0 %      | 37246 |       |       |       |       | 1  |

## LB119548 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Apr 2022 12:19:49

| Sample ID       | Extended ID | μ Abs. | Conc.  | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|--------|--------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50          | 443    | -      | 0.0000   | 7470A  | PPB   | 06 Apr 2022 12:21:27 | S    | Std  |
| 0.2 - 1         | 50-2        | 1009   | -      | 0.2000   | 7470A  | PPB   | 06 Apr 2022 12:23:44 | S    | Std  |
| 2.5 - 1         | 52.5        | 8880   | -      | 2.5000   | 7470A  | PPB   | 06 Apr 2022 12:26:01 | S    | Std  |
| 5.0 - 1         | 55          | 18675  | -      | 5.0000   | 7470A  | PPB   | 06 Apr 2022 12:28:17 | S    | Std  |
| 7.5 - 1         | 57.5        | 27553  | -      | 7.5000   | 7470A  | PPB   | 06 Apr 2022 12:30:34 | S    | Std  |
| 10.0 - 1        | 510         | 37246  | -      | 10.0000  | 7470A  | PPB   | 06 Apr 2022 12:32:53 | S    | Std  |
| ICV65 - 1       | ICV65       | 15021  | 4.0335 | -        | 7470A  | PPB   | 06 Apr 2022 12:38:07 | U    | SMPL |
| ICB65 - 1       | ICB65       | 299    | 0.0379 | -        | 7470A  | PPB   | 06 Apr 2022 12:40:23 | U    | SMPL |
| CCV96 - 1       | CCV96       | 18345  | 4.9357 | -        | 7470A  | PPB   | 06 Apr 2022 12:42:40 | U    | SMPL |
| CCB96 - 1       | CCB96       | 309    | 0.0406 | -        | 7470A  | PPB   | 06 Apr 2022 12:44:55 | U    | SMPL |
| CRA - 1         | CRA         | 891    | 0.1986 | -        | 7470A  | PPB   | 06 Apr 2022 12:47:12 | U    | SMPL |
| HighStd - 1     | HighStd     | 36931  | 9.9801 | -        | 7470A  | PPB   | 06 Apr 2022 12:49:27 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 25713  | 6.9354 | -        | 7470A  | PPB   | 06 Apr 2022 12:51:42 | U    | SMPL |
| PB143926BL - 1  | PBW         | 383    | 0.0607 | -        | 7470A  | PPB   | 06 Apr 2022 12:56:49 | U    | SMPL |
| PB143926BS - 1  | LCSW        | 15104  | 4.0561 | -        | 7470A  | PPB   | 06 Apr 2022 12:59:05 | U    | SMPL |
| N2253-06 - 1    | 2822S       | 4196   | 1.0956 | -        | 7470A  | PPB   | 06 Apr 2022 13:01:21 | U    | SMPL |
| N2269-01 - 1    | TP-6        | 1691   | 0.4157 | -        | 7470A  | PPB   | 06 Apr 2022 13:03:39 | U    | SMPL |
| N2269-01DUP - 1 | TP-6DUP     | 1519   | 0.3690 | -        | 7470A  | PPB   | 06 Apr 2022 13:05:57 | U    | SMPL |
| N2269-01MS - 1  | TP-6MS      | 18082  | 4.8643 | -        | 7470A  | PPB   | 06 Apr 2022 13:08:15 | U    | SMPL |
| N2269-01MSD - 1 | TP-6MSD     | 18470  | 4.9696 | -        | 7470A  | PPB   | 06 Apr 2022 13:10:30 | U    | SMPL |
| CCV97 - 1       | CCV97       | 18087  | 4.8657 | -        | 7470A  | PPB   | 06 Apr 2022 13:12:46 | U    | SMPL |
| CCB97 - 1       | CCB97       | 362    | 0.0550 | -        | 7470A  | PPB   | 06 Apr 2022 13:15:03 | U    | SMPL |
| N2269-02 - 1    | TP-5        | 1084   | 0.2509 | -        | 7470A  | PPB   | 06 Apr 2022 13:17:20 | U    | SMPL |
| N2269-03 - 1    | LP-2        | 1091   | 0.2528 | -        | 7470A  | PPB   | 06 Apr 2022 13:19:35 | U    | SMPL |
| N2269-01LX5 - 1 |             | 568    | 0.1109 | -        | 7470A  | PPB   | 06 Apr 2022 13:21:50 | U    | SMPL |
| N2269-01A - 1   |             | 16626  | 4.4691 | -        | 7470A  | PPB   | 06 Apr 2022 13:24:06 | U    | SMPL |
| CCV98 - 1       | CCV98       | 18097  | 4.8684 | -        | 7470A  | PPB   | 06 Apr 2022 13:26:21 | U    | SMPL |
| CCB98 - 1       | CCB98       | 179    | 0.0053 | -        | 7470A  | PPB   | 06 Apr 2022 13:28:39 | U    | SMPL |