



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119555

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	04/06/2022	16:45
2	CAL2	1	Water	NA	NA	20.5	7.01	04/06/2022	16:46
3	CAL3	1	Water	NA	NA	20.5	10.03	04/06/2022	16:47
4	ICV	1	Water	NA	NA	20.6	6.99	04/06/2022	16:50
5	CCV1	1	Water	NA	NA	20.6	2.01	04/06/2022	16:52
6	N2288-01	1	Water	NA	NA	20.1	5.91	04/06/2022	16:55
7	N2302-01	1	Water	NA	NA	20.6	7.16	04/06/2022	17:00
8	N2302-01DUP	1	Water	NA	NA	20.7	7.17	04/06/2022	17:01
9	CCV2	1	Water	NA	NA	20.6	12.03	04/06/2022	17:04

119555

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2302 WorkList ID : 158438 Department : Wet-Chemistry Date : 04-06-2022 16:38:05

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
04/06/2022	Water	N2288-01	Corrosivity	Cool 4 deg C	PSEG03	O11	TOTE	04/06/2022	9040C
04/06/2022	Water	N2302-01	Corrosivity	Cool 4 deg C	PSEG03	O11	1-3-TOTES	04/06/2022	9040C

Date/Time 04/06/22 16:40
Raw Sample Received by: 201601
Raw Sample Relinquished by: L-C (Sun)

Date/Time 04/08/22 17:45
Raw Sample Received by: L-C (Sun)
Raw Sample Relinquished by: SP 1601