



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119560

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	04/07/2022	09:33
2	CAL2	1	Water	NA	NA	20.6	7.00	04/07/2022	09:34
3	CAL3	1	Water	NA	NA	20.6	10.03	04/07/2022	09:35
4	ICV	1	Water	NA	NA	20.6	7.01	04/07/2022	09:36
5	CCV1	1	Water	NA	NA	20.6	2.01	04/07/2022	09:39
6	N2292-04	1	Water	NA	NA	20.6	6.47	04/07/2022	09:40
7	N2292-05	1	Water	NA	NA	20.7	7.12	04/07/2022	09:41
8	N2292-06	1	Water	NA	NA	20.6	9.20	04/07/2022	09:42
9	N2294-01	1	Water	NA	NA	20.5	7.64	04/07/2022	09:45
10	N2298-03	1	Water	NA	NA	20.8	7.44	04/07/2022	09:46
11	N2298-05	1	Water	NA	NA	20.7	6.96	04/07/2022	09:47
12	N2306-03	1	Water	NA	NA	20.8	6.98	04/07/2022	09:50
13	N2306-04	1	Water	NA	NA	20.7	7.14	04/07/2022	09:52
14	N2306-05	1	Water	NA	NA	21.2	6.86	04/07/2022	09:55
15	N2306-06	1	Water	NA	NA	21.1	10.43	04/07/2022	10:00
16	CCV2	1	Water	NA	NA	20.7	12.02	04/07/2022	10:01
17	N2306-07	1	Water	NA	NA	21.0	7.24	04/07/2022	10:03
18	N2306-08	1	Water	NA	NA	21.1	7.01	04/07/2022	10:05
19	N2306-09	1	Water	NA	NA	21.3	8.68	04/07/2022	10:06
20	N2306-09DUP	1	Water	NA	NA	21.4	8.69	04/07/2022	10:07

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.4	2.02	04/07/2022	10:10

119560

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2306 WorkList ID : 158454 Department : Wet-Chemistry Date : 04-07-2022 09:25:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
04/06/2022	Water	N2292-04	pH	Cool 4 deg C	PSEG03	O11	2282	04/06/2022	9040C
04/06/2022	Water	N2292-05	pH	Cool 4 deg C	PSEG03	O11	2283	04/06/2022	9040C
04/06/2022	Water	N2292-06	pH	Cool 4 deg C	PSEG03	O11	2284	04/06/2022	9040C
04/06/2022	Water	N2294-01	pH	Cool 4 deg C	PSEG04	O11	402	04/06/2022	9040C
04/06/2022	Water	N2298-03	pH	Cool 4 deg C	PSEG03	O11	FMI-126	04/06/2022	9040C
04/06/2022	Water	N2298-05	pH	Cool 4 deg C	PSEG03	O11	TOTE	04/06/2022	9040C
04/06/2022	Water	N2306-03	pH	Cool 4 deg C	PSEG03	O11	2421	04/06/2022	9040C
04/06/2022	Water	N2306-04	pH	Cool 4 deg C	PSEG03	O11	2422	04/06/2022	9040C
04/06/2022	Water	N2306-05	pH	Cool 4 deg C	PSEG03	O11	2383	04/06/2022	9040C
04/06/2022	Water	N2306-06	pH	Cool 4 deg C	PSEG03	O11	2384	04/06/2022	9040C
04/06/2022	Water	N2306-07	pH	Cool 4 deg C	PSEG03	O11	2385	04/06/2022	9040C
04/06/2022	Water	N2306-08	pH	Cool 4 deg C	PSEG03	O11	2386	04/06/2022	9040C
04/06/2022	Water	N2306-09	pH	Cool 4 deg C	PSEG03	O11	2431	04/06/2022	9040C

04/07/22 091.50

04/07/22 11.00

Date/Time
Raw Sample Received by: JP (GCI)
Raw Sample Relinquished by: CD SM

Date/Time
Raw Sample Received by: JP SM
Raw Sample Relinquished by: JP (GCI)