



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119561

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	04/07/2022	09:33
2	CAL2	1	Water	NA	NA	20.6	7.00	04/07/2022	09:34
3	CAL3	1	Water	NA	NA	20.6	10.03	04/07/2022	09:35
4	ICV	1	Water	NA	NA	20.6	7.01	04/07/2022	09:36
5	CCV1	1	Water	NA	NA	20.6	2.01	04/07/2022	09:36
6	N2292-01	1	Solid	20.02	20	20.3	4.51	04/07/2022	10:15
7	N2292-02	1	Solid	20.01	20	20.4	4.56	04/07/2022	10:16
8	N2292-03	1	Solid	20.02	20	20.5	4.63	04/07/2022	10:19
9	N2292-03DUP	1	Solid	20.03	20	20.6	4.64	04/07/2022	10:20
10	CCV2	1	Water	NA	NA	20.7	12.03	04/07/2022	10:22

119561

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2292 S WorkList ID : 158455 Department : Wet-Chemistry Date : 04-07-2022 09:26:20

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Raw Sample Storage Location		
04/06/2022	Solid	N2292-01	pH <i>✓</i>	Cool 4 deg C	PSEG03	O11 2291	04/06/2022	9045D
04/06/2022	Solid	N2292-02	pH <i>✓</i>	Cool 4 deg C	PSEG03	O11 2286	04/06/2022	9045D
04/06/2022	Solid	N2292-03	pH <i>✓</i>	Cool 4 deg C	PSEG03	O11 2287	04/06/2022	9045D

Date/Time 04/07/22 09:13 F
Raw Sample Received by: SP (WC)
Raw Sample Relinquished by: SP Ssr

Date/Time 04/07/22 11:15
Raw Sample Received by: SP Ssr
Raw Sample Relinquished by: SP (WC)