



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB119565
Analysis Date: 04/07/2022
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 04/07/2022
Extraction IN Time: 08:30
Extraction OUT Time: 09:30
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB119565BL	LB119565BL	WATER	1.3	1000	100	2.9258	2.9258	3.01	2.9258	2.9258	0.0000	0
2	LB119565BS	LB119565BS	WATER	1.3	1000	100	2.8743	2.8743	3.02	2.8911	2.8911	0.0168	16.8
3	LB119565BSD	LB119565BSD	WATER	1.3	1000	100	2.9366	2.9366	3.00	2.9533	2.9533	0.0167	16.7
4	N2173-01	GRAB	WATER	1.6	1000	100	2.9101	2.9101	3.03	2.9254	2.9254	0.0153	15.3
5	N2285-01	GRAB	WATER	1.6	1000	100	2.9975	2.9975	3.04	3.0468	3.0468	0.0493	49.3
6	N2286-08	8G	WATER	1.6	1000	100	2.9926	2.9926	3.03	2.9928	2.9928	0.0002	0.2
7	N2294-01	402	WATER	1.3	1000	100	2.9868	2.9868	3.04	2.9870	2.9870	0.0002	0.2



Chemical Name	Chemical Lot #
HEXANE	W2885
pH Paper 0-14	M4909
Sodium Sulfate	EP2222
1:1 HCL	WP95586
Silica Gel	W2839
Sand	E2865

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP95587
LCSWD	5.00 ML	WP95588
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0019</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>10:30</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>10:00</u>		
Bal Check Time:	<u>08:45</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>11:00</u>
			Out Time1	<u>10:01</u>		

0.0020 gram Balance:	<u>0.0018</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>12:02</u>
1.0000 gram Balance:	<u>1.0004</u>	(0.9950-1.0050)	In Time2:	<u>11:30</u>		
Bal Check Time:	<u>12:45</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>12:37</u>
			Out Time2:	<u>12:00</u>		

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH N2286

WorkList ID : 158442

Department : Wet-Chemistry

Date : 04-07-2022 08:13:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/12/2022	Water	N2173-01	TPH	Conc H2SO4 to	ARAM01	N22	GRAB	03/30/2022	1664A
04/19/2022	Water	N2285-01	TPH	Conc H2SO4 to	ARAM01	N21	GRAB	04/06/2022	1664A
04/13/2022	Water	N2286-08	TPH	Conc H2SO4 to	M&MM01	O11	8G	04/06/2022	1664A
04/19/2022	Water	N2294-01	TPH	Conc H2SO4 to	PSEG04	O11	402	04/06/2022	1664A

Date/Time 04/07/22 08:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/07/22
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]