



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119572

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP96225
calibration std. phosphate 0.5 ppm	WP96224
calibration std. phosphate 0.3 ppm	WP96223
calibration std. phosphate 0.1 ppm	WP96222
calibration std. phosphate 0.05 ppm	WP96221
calibration std. 0 ppm	WP96220
phosphate CCV std.	WP96226
Combined reagent	WP96229
Phenolphthalein indicator	WP93767
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP96227

Intercept: -0.0015

Slope: 0.6517

Regression: 0.999923

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		04/07/2022	14:45
2	CAL2	0.05	1	50	50	0.033	0.053	-6	04/07/2022	14:45
3	CAL3	0.10	1	50	50	0.065	0.102	-2	04/07/2022	14:46
4	CAL4	0.30	1	50	50	0.188	0.291	3	04/07/2022	14:46
5	CAL5	0.50	1	50	50	0.324	0.499	0.2	04/07/2022	14:47
6	CAL6	1.00	1	50	50	0.652	1.003	-0.3	04/07/2022	14:47



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119572

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.329	0.507	04/07/2022	14:48
2	ICB		1	50	50	0.000	0.002	04/07/2022	14:48
3	CCV1	0.50	1	50	50	0.325	0.501	04/07/2022	14:49
4	CCB1		1	50	50	0.001	0.004	04/07/2022	14:49
5	PB143950BL		1	50	50	0.002	0.005	04/07/2022	14:50
6	PB143950BS	0.50	1	50	50	0.333	0.513	04/07/2022	14:50
7	N2284-01		1	50	50	0.011	0.019	04/07/2022	14:51
8	N2284-02		1	50	50	0.027	0.044	04/07/2022	14:51
9	N2284-04		1	50	50	0.240	0.371	04/07/2022	14:52
10	N2284-05		1	50	50	0.233	0.360	04/07/2022	14:52
11	N2284-06		1	50	50	0.232	0.358	04/07/2022	14:53
12	N2284-06DUP		1	50	50	0.231	0.357	04/07/2022	14:53
13	N2284-06MS	0.50	1	50	50	0.582	0.895	04/07/2022	14:54
14	N2284-06MSD	0.50	1	50	50	0.581	0.894	04/07/2022	14:54
15	CCV2	0.50	1	50	50	0.327	0.504	04/07/2022	14:55
16	CCB2		1	50	50	0.000	0.002	04/07/2022	14:55
17	N2284-07		1	50	50	0.230	0.355	04/07/2022	14:56
18	N2286-05		1	50	50	1.079	1.658	04/07/2022	14:56
19	N2286-05		5	50	50	0.604	0.929	04/07/2022	14:57
20	CCV3	0.50	1	50	50	0.326	0.503	04/07/2022	14:57
21	CCB3		1	50	50	0.001	0.004	04/07/2022	14:58