



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119576

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/07/2022	16:36
2	CAL2	1	Water	NA	NA	20.7	7.00	04/07/2022	16:37
3	CAL3	1	Water	NA	NA	20.7	10.03	04/07/2022	16:39
4	ICV	1	Water	NA	NA	20.5	6.99	04/07/2022	16:40
5	CCV1	1	Water	NA	NA	20.7	2.01	04/07/2022	16:42
6	N2307-01	1	Water	NA	NA	20.1	5.79	04/07/2022	16:45
7	N2313-08	1	Water	NA	NA	20.1	6.64	04/07/2022	16:46
8	N2325-01	1	Water	NA	NA	20.3	7.16	04/07/2022	16:50
9	N2325-03	1	Water	NA	NA	20.6	7.13	04/07/2022	16:52
10	N2325-05	1	Water	NA	NA	20.3	7.00	04/07/2022	17:02
11	N2327-01	1	Water	NA	NA	20.3	6.49	04/07/2022	17:05
12	N2327-01DUP	1	Water	NA	NA	20.4	6.50	04/07/2022	17:06
13	CCV2	1	Water	NA	NA	20.7	12.03	04/07/2022	17:15

LB 119576

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2327 WorkList ID : 158492 Department : Wet-Chemistry Date : 04-07-2022 16:24:51

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
04/07/2022	Water	N2307-01	pH <i>A</i>	Cool 4 deg C	PSEG03	O11 F05421	04/07/2022	9040C
04/07/2022	Water	N2313-08	pH	Cool 4 deg C	PSEG03	O11 DRUM-33604	04/07/2022	9040C
04/07/2022	Water	N2325-01	pH	Cool 4 deg C	PSEG04	O11 DSN002	04/07/2022	9040C
04/07/2022	Water	N2325-03	pH	Cool 4 deg C	PSEG04	O11 DSN001	04/07/2022	9040C
04/07/2022	Water	N2325-05	pH	Cool 4 deg C	PSEG04	O11 RWS003	04/07/2022	9040C
04/07/2022	Water	N2327-01	pH <i>A</i>	Cool 4 deg C	PSEG03	O11 A4128	04/07/2022	9040C

Date/Time
04/07/22 16:30

Raw Sample Received by:
JR (SM)

Raw Sample Relinquished by:
JR (SM)

Date/Time
04/07/22 17:40

Raw Sample Received by:
JR (SM)

Raw Sample Relinquished by:
JR (SM)