



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119580

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	27.00	80.6	81.2	04/08/2022	08:30
2	N2288-01		1	10	100.00	>212.0	>212.0	04/08/2022	08:46
3	N2288-01DUP		1	10	100.00	>212.0	>212.0	04/08/2022	09:03
4	N2292-07		1	9	100.00	>212.0	>212.0	04/08/2022	09:20
5	N2292-09		1	9	100.00	>212.0	>212.0	04/08/2022	09:46
6	N2292-11		1	10	100.00	>212.0	>212.0	04/08/2022	10:01
7	N2292-13		1	11	100.00	>212.0	>212.0	04/08/2022	10:19
8	N2298-03		1	10	100.00	>212.0	>212.0	04/08/2022	10:35
9	N2298-05		1	10	100.00	>212.0	>212.0	04/08/2022	10:52
10	N2302-01		1	11	100.00	>212.0	>212.0	04/08/2022	11:10
11	N2306-09		1	10	100.00	>212.0	>212.0	04/08/2022	11:26
12	N2306-10		1	10	100.00	>212.0	>212.0	04/08/2022	11:42
13	N2306-12		1	9	100.00	>212.0	>212.0	04/08/2022	11:59
14	N2306-14		1	9	100.00	>212.0	>212.0	04/08/2022	12:16
15	N2307-01		1	10	100.00	>212.0	>212.0	04/08/2022	12:33
16	N2313-08		1	10	100.00	>212.0	>212.0	04/08/2022	13:29
17	N2327-01		1	10	100.00	>212.0	>212.0	04/08/2022	13:45
18	N2327-01DUP		1	10	100.00	>212.0	>212.0	04/08/2022	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 040722 WorkList ID : 158439 Department : Wet-Chemistry Date : 04-06-2022 17:13:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/13/2022	Water	N2288-01	Flash Point ✓	Cool 4 deg C	PSEG03	O11	TOTE	04/06/2022	1010B
04/11/2022	Water	N2292-07	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2282-84	04/06/2022	1010B
04/11/2022	Water	N2292-09	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2266	04/06/2022	1010B
04/11/2022	Water	N2292-11	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2256	04/06/2022	1010B
04/11/2022	Water	N2292-13	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2238	04/06/2022	1010B
04/11/2022	Water	N2298-03	Flash Point ✓	Cool 4 deg C	PSEG03	O11	FMI-126	04/06/2022	1010B
04/11/2022	Water	N2298-05	Flash Point ✓	Cool 4 deg C	PSEG03	O11	TOTE	04/06/2022	1010B
04/19/2022	Water	N2302-01	Flash Point ✓	Cool 4 deg C	PSEG03	O11	1-3-TOTES	04/06/2022	1010B
04/11/2022	Water	N2306-09	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2431	04/06/2022	1010B
04/11/2022	Water	N2306-10	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2433	04/06/2022	1010B
04/11/2022	Water	N2306-12	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2421-22	04/06/2022	1010B
04/11/2022	Water	N2306-14	Flash Point ✓	Cool 4 deg C	PSEG03	O11	2383-2386	04/06/2022	1010B
04/12/2022	Water	N2307-01	Flash Point ✓	Cool 4 deg C	PSEG03	O11	F05421	04/07/2022	1010B
04/14/2022	Water	N2313-08	Flash Point ✓	Cool 4 deg C	PSEG03	O11	DRUM-33604	04/07/2022	1010B
04/12/2022	Water	N2327-01	Flash Point ✓	Cool 4 deg C	PSEG03	O11	A4128	04/07/2022	1010B

Date/Time 4/8/22 8:15
Raw Sample Received by: AP (WC)
Raw Sample Relinquished by: AP SR

Date/Time 4/8/22 14:30
Raw Sample Received by: AP SR
Raw Sample Relinquished by: AP (WC)