


LB119583

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1218e-004

C= -3.1251e-002

Rho= 0.9993693

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.083	0.083	539	0.000	539					-
0.2	0.200	0.204	0.004	1110	0.0 %	1110					2
2.5	2.500	2.409	-0.091	11502	0.0 %	11502					-4
5.0	5.000	5.102	0.102	24194	0.0 %	24194					2
7.5	7.500	7.260	-0.240	34365	0.0 %	34365					-3
10.0	10.000	10.141	0.141	47943	0.0 %	47943					1

LB119583 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 08 Apr 2022 10:54:29					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	539	-	0.0000	7471B	PPB	08 Apr 2022 13:30:16	S	Std
0.2 - 1	50-2	1110	-	0.2000	7471B	PPB	08 Apr 2022 13:32:34	S	Std
2.5 - 1	52-5	11502	-	2.5000	7471B	PPB	08 Apr 2022 13:34:50	S	Std
5.0 - 1	55	24194	-	5.0000	7471B	PPB	08 Apr 2022 13:37:07	S	Std
7.5 - 1	57-1	34365	-	7.5000	7471B	PPB	08 Apr 2022 13:39:23	S	Std
10.0 - 1	512	47943	-	10.0000	7471B	PPB	08 Apr 2022 13:44:30	S	Std
ICV67 - 1	ICV67	18084	3.8057	-	7471B	PPB	08 Apr 2022 13:47:20	U	SMPL
ICB67 - 1	ICB67	79	-0.0145	-	7471B	PPB	08 Apr 2022 13:49:35	U	SMPL
CCV06 - 1	CCV06	23610	4.9782	-	7471B	PPB	08 Apr 2022 13:51:51	U	SMPL
CCB06 - 1	CCB06	57	-0.0192	-	7471B	PPB	08 Apr 2022 13:54:08	U	SMPL
CRA - 1	CRA	1058	0.1932	-	7471B	PPB	08 Apr 2022 13:56:25	U	SMPL
HighStd - 1	HighStd	47033	9.9480	-	7471B	PPB	08 Apr 2022 13:58:40	U	SMPL
ChkStd - 1	ChkStd	31865	6.7297	-	7471B	PPB	08 Apr 2022 14:00:56	U	SMPL
PB143930BL - 1	PBS	363	0.0458	-	7471B	PPB	08 Apr 2022 14:03:14	U	SMPL
PB143930BS - 1	LCSS	18348	3.8618	-	7471B	PPB	08 Apr 2022 14:05:31	U	SMPL
N2252-01 - 1	OK-01-040422	4126	0.8442	-	7471B	PPB	08 Apr 2022 14:07:47	U	SMPL
N2252-01DUP - 1	OK-01-040422DUP	4253	0.8711	-	7471B	PPB	08 Apr 2022 14:10:19	U	SMPL
N2252-01MS - 1	OK-01-040422MS	21701	4.5732	-	7471B	PPB	08 Apr 2022 14:12:36	U	SMPL
N2252-01MSD - 1	OK-01-040422MSD	22075	4.6525	-	7471B	PPB	08 Apr 2022 14:14:53	U	SMPL
N2253-01 - 1	101221	2201	0.4357	-	7471B	PPB	08 Apr 2022 14:17:09	U	SMPL
CCV07 - 1	CCV07	23906	5.0410	-	7471B	PPB	08 Apr 2022 14:19:24	U	SMPL
CCB07 - 1	CCB07	0	-0.0313	-	7471B	PPB	08 Apr 2022 14:21:39	U	SMPL
N2253-02 - 1	31292	3923	0.8011	-	7471B	PPB	08 Apr 2022 14:23:54	U	SMPL
N2253-03 - 1	31396	495	0.0738	-	7471B	PPB	08 Apr 2022 14:26:09	U	SMPL
N2254-01 - 1	WC-14A-1A-4-1	65276	13.8188	-	7471B	PPB	08 Apr 2022 14:30:59	U	SMPL
N2254-03 - 1	WC-14A-1A-4-2	6494	1.3466	-	7471B	PPB	08 Apr 2022 14:33:14	U	SMPL
N2254-05 - 1	WC-14A-1A-4-3	4720	0.9702	-	7471B	PPB	08 Apr 2022 14:35:31	U	SMPL
N2256-01 - 1	SU-02-040422	1430	0.2722	-	7471B	PPB	08 Apr 2022 14:37:48	U	SMPL
N2256-03 - 1	SU-03-040422	2953	0.5953	-	7471B	PPB	08 Apr 2022 14:40:04	U	SMPL
N2257-01 - 1	TP-1	38154	8.0641	-	7471B	PPB	08 Apr 2022 14:42:21	U	SMPL
N2257-03 - 1	TP-2	22578	4.7593	-	7471B	PPB	08 Apr 2022 14:44:37	U	SMPL
N2257-05 - 1	TP-3	15853	3.3324	-	7471B	PPB	08 Apr 2022 14:46:54	U	SMPL
CCV08 - 1	CCV08	23486	4.9519	-	7471B	PPB	08 Apr 2022 14:49:09	U	SMPL
CCB08 - 1	CCB08	298	0.0320	-	7471B	PPB	08 Apr 2022 14:51:24	U	SMPL
N2257-07 - 1	TP-4	12641	2.6509	-	7471B	PPB	08 Apr 2022 14:53:40	U	SMPL
N2257-09 - 1	TP-5	15697	3.2993	-	7471B	PPB	08 Apr 2022 14:55:55	U	SMPL
N2257-11 - 1	TP-6	22690	4.7830	-	7471B	PPB	08 Apr 2022 14:58:11	U	SMPL
PB143931BL - 1	PBS	422	0.0583	-	7471B	PPB	08 Apr 2022 15:00:26	U	SMPL
PB143931BS - 1	LCSS	17603	3.7037	-	7471B	PPB	08 Apr 2022 15:02:42	U	SMPL
N2263-01 - 1	LINDEN-NORTH	4761	0.9789	-	7471B	PPB	08 Apr 2022 15:04:58	U	SMPL
N2267-01 - 1	LINDEN-SOUTH	50334	10.6484	-	7471B	PPB	08 Apr 2022 15:07:14	U	SMPL
N2270-01 - 1	FARMINGDALE-CLAY-1	5556	1.1476	-	7471B	PPB	08 Apr 2022 15:09:31	U	SMPL
N2272-03 - 1	TP-1	1145	0.2117	-	7471B	PPB	08 Apr 2022 15:11:49	U	SMPL
N2272-07 - 1	TP-3	1934	0.3791	-	7471B	PPB	08 Apr 2022 15:14:04	U	SMPL
CCV09 - 1	CCV09	23416	4.9371	-	7471B	PPB	08 Apr 2022 15:16:20	U	SMPL
CCB09 - 1	CCB09	364	0.0460	-	7471B	PPB	08 Apr 2022 15:18:35	U	SMPL
N2275-01 - 1	EO-03-040522	3162	0.6396	-	7471B	PPB	08 Apr 2022 15:20:51	U	SMPL
N2277-01 - 1	PF-1	821	0.1429	-	7471B	PPB	08 Apr 2022 15:23:06	U	SMPL
N2278-01 - 1	OR-03-040522	2080	0.4101	-	7471B	PPB	08 Apr 2022 15:25:22	U	SMPL
N2278-01DUP - 1	OR-03-040522DUP	2347	0.4667	-	7471B	PPB	08 Apr 2022 15:27:37	U	SMPL
N2278-01MS - 1	OR-03-040522MS	21085	4.4425	-	7471B	PPB	08 Apr 2022 15:29:52	U	SMPL
N2278-01MSD - 1	OR-03-040522MSD	20269	4.2693	-	7471B	PPB	08 Apr 2022 15:32:09	U	SMPL
N2252-01LX5 - 1		648	0.1062	-	7471B	PPB	08 Apr 2022 15:34:25	U	SMPL
N2252-01A - 1		21324	4.4932	-	7471B	PPB	08 Apr 2022 15:36:41	U	SMPL
N2278-01LX5 - 1		321	0.0369	-	7471B	PPB	08 Apr 2022 15:38:58	U	SMPL
N2278-01A - 1		21513	4.5333	-	7471B	PPB	08 Apr 2022 15:41:15	U	SMPL
CCV10 - 1	CCV10	23642	4.9850	-	7471B	PPB	08 Apr 2022 15:43:30	U	SMPL
CCB10 - 1	CCB10	253	0.0224	-	7471B	PPB	08 Apr 2022 15:45:46	U	SMPL
N2254-01DLX2 - 1		32698	6.9065	-	7471B	PPB	08 Apr 2022 16:10:24	U	SMPL
N2267-01DLX2 - 1		25872	5.4582	-	7471B	PPB	08 Apr 2022 16:12:40	U	SMPL
CCV11 - 1	CCV11	25571	5.3943	-	7471B	PPB	08 Apr 2022 16:14:56	U	SMPL
CCB11 - 1	CCB11	64	-0.0177	-	7471B	PPB	08 Apr 2022 16:17:13	U	SMPL