



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119597

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	04/08/2022	12:35
2	CAL2	1	Water	NA	NA	20.4	7.00	04/08/2022	12:36
3	CAL3	1	Water	NA	NA	20.4	10.03	04/08/2022	12:37
4	ICV	1	Water	NA	NA	20.5	6.99	04/08/2022	12:39
5	CCV1	1	Water	NA	NA	20.5	2.01	04/08/2022	12:40
6	N2341-02	1	Water	NA	NA	20.2	7.22	04/08/2022	12:45
7	N2341-02DUP	1	Water	NA	NA	20.3	7.23	04/08/2022	12:46
8	CCV2	1	Water	NA	NA	20.5	12.02	04/08/2022	12:49

UP119591f

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH N2341	WorkList ID :	158519	Department :	Wet-Chemistry	Date :	04-08-2022 12:29:58		
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/07/2022	Water	N2341-02	pH	Cool 4 deg C	TEIR06	O11	TT-040-IDWGW-20220407	04/07/2022	9040C

Date/Time 04/08/22 12:30
Raw Sample Received by: J.L. (SM)
Raw Sample Relinquished by: J.L. (SM)

Date/Time 04/08/22 12:00
Raw Sample Received by: J.L. (SM)
Raw Sample Relinquished by: J.L. (SM)