


LB119602

7470A

INSTRUMENT ID: CV4

Linear



A= 0.0000e+000

B= 2.9213e-004

C= 1.4988e-002

Rho= 0.9989794

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.0	0.000	0.009	0.009	-19	0.000	-19					1
0.2	0.200	0.205	0.005	650	0.0 %	650					2
2.5	2.500	2.292	-0.208	7793	0.0 %	7793					8
5.0	5.000	5.238	0.238	17880	0.0 %	17880					5
7.5	7.500	7.659	0.159	26166	0.0 %	26166					2
10.0	10.000	9.797	-0.203	33486	0.0 %	33486					2

LB119602 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Apr 2022 06:25:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7470A	PPB	09 Apr 2022 06:30:08	S	Std
0.2 - 1	50.2	650	-	0.2000	7470A	PPB	09 Apr 2022 06:32:24	S	Std
2.5 - 1	52.5	7793	-	2.5000	7470A	PPB	09 Apr 2022 06:34:41	S	Std
5.0 - 1	55	17880	-	5.0000	7470A	PPB	09 Apr 2022 06:36:57	S	Std
7.5 - 1	57.5	26166	-	7.5000	7470A	PPB	09 Apr 2022 06:39:14	S	Std
10.0 - 1	510	33486	-	10.0000	7470A	PPB	09 Apr 2022 06:41:32	S	Std
ICV68 - 1	ICV68	12628	3.7040	-7470A	PPB	09 Apr 2022 06:44:21	U	SMPL	
ICB68 - 1	ICB68	-100	-0.0142	-7470A	PPB	09 Apr 2022 06:46:36	U	SMPL	
CCV12 - 1	CCV12	16729	4.9020	-7470A	PPB	09 Apr 2022 06:48:51	U	SMPL	
CCB12 - 1	CCB12	-113	-0.0180	-7470A	PPB	09 Apr 2022 06:51:05	U	SMPL	
CRA - 1	CRA	650	0.2049	-7470A	PPB	09 Apr 2022 06:53:20	U	SMPL	
HighStd - 1	HighStd	33464	9.7907	-7470A	PPB	09 Apr 2022 06:55:35	U	SMPL	
ChkStd - 1	ChkStd	23602	6.9098	-7470A	PPB	09 Apr 2022 06:57:50	U	SMPL	
PB143944BL - 1	PBW	-181	-0.0379	-7470A	PPB	09 Apr 2022 07:00:09	U	SMPL	
PB143944BS - 1	LCSW	14229	4.1717	-7470A	PPB	09 Apr 2022 07:02:26	U	SMPL	
N2253-04 - 1	2822A	2373	0.7082	-7470A	PPB	09 Apr 2022 07:04:42	U	SMPL	
N2254-02 - 1	WC-14A-1A-4-1	-9	0.0124	-7470A	PPB	09 Apr 2022 07:09:20	U	SMPL	
N2254-04 - 1	WC-14A-1A-4-2	74	0.0366	-7470A	PPB	09 Apr 2022 07:11:38	U	SMPL	
N2254-06 - 1	WC-14A-1A-4-3	113	0.0480	-7470A	PPB	09 Apr 2022 07:13:55	U	SMPL	
N2255-01 - 1	31496	843	0.2613	-7470A	PPB	09 Apr 2022 07:16:10	U	SMPL	
CCV13 - 1	CCV13	15445	4.5269	-7470A	PPB	09 Apr 2022 07:18:26	U	SMPL	
CCB13 - 1	CCB13	-61	-0.0028	-7470A	PPB	09 Apr 2022 07:20:41	U	SMPL	
N2267-02 - 1	LINDEN-SOUTH	-15	0.0106	-7470A	PPB	09 Apr 2022 07:22:56	U	SMPL	
N2272-04 - 1	TP-1	6	0.0167	-7470A	PPB	09 Apr 2022 07:25:11	U	SMPL	
N2272-08 - 1	TP-3	6	0.0167	-7470A	PPB	09 Apr 2022 07:27:27	U	SMPL	
N2272-08DUP - 1	TP-3DUP	23	0.0217	-7470A	PPB	09 Apr 2022 07:29:42	U	SMPL	
N2272-08MS - 1	TP-3MS	12485	3.6622	-7470A	PPB	09 Apr 2022 07:31:58	U	SMPL	
N2272-08MSD - 1	TP-3MSD	13167	3.8614	-7470A	PPB	09 Apr 2022 07:34:14	U	SMPL	
PB143963BL - 1	PBW	-120	-0.0201	-7470A	PPB	09 Apr 2022 07:36:33	U	SMPL	
PB143963BS - 1	LCSW	12795	3.7528	-7470A	PPB	09 Apr 2022 07:38:49	U	SMPL	
N2292-07 - 1	2282-84	1527	0.4611	-7470A	PPB	09 Apr 2022 07:41:06	U	SMPL	
N2292-09 - 1	2266	0	0.0150	-7470A	PPB	09 Apr 2022 07:43:21	U	SMPL	
CCV14 - 1	CCV14	16133	4.7279	-7470A	PPB	09 Apr 2022 07:45:37	U	SMPL	
CCB14 - 1	CCB14	-76	-0.0072	-7470A	PPB	09 Apr 2022 07:47:52	U	SMPL	
N2292-11 - 1	2256	763	0.2379	-7470A	PPB	09 Apr 2022 07:50:08	U	SMPL	
N2292-13 - 1	2238	178	0.0670	-7470A	PPB	09 Apr 2022 07:52:23	U	SMPL	
N2298-03 - 1	FMI-126	-13	0.0112	-7470A	PPB	09 Apr 2022 07:54:39	U	SMPL	
N2298-05 - 1	TOTE	1027	0.3150	-7470A	PPB	09 Apr 2022 07:56:55	U	SMPL	
N2303-01 - 1	MW-109	18	0.0202	-7470A	PPB	09 Apr 2022 07:59:10	U	SMPL	
N2303-02 - 1	MW-104	0	0.0150	-7470A	PPB	09 Apr 2022 08:01:26	U	SMPL	
N2303-03 - 1	MW-101	0	0.0150	-7470A	PPB	09 Apr 2022 08:03:42	U	SMPL	
N2303-04 - 1	MW-105	-2	0.0144	-7470A	PPB	09 Apr 2022 08:06:05	U	SMPL	
N2303-05 - 1	MW-105	34	0.0249	-7470A	PPB	09 Apr 2022 08:08:22	U	SMPL	
N2303-06 - 1	MW-103	400	0.1318	-7470A	PPB	09 Apr 2022 08:10:37	U	SMPL	
CCV15 - 1	CCV15	16305	4.7781	-7470A	PPB	09 Apr 2022 08:12:53	U	SMPL	
CCB15 - 1	CCB15	150	0.0588	-7470A	PPB	09 Apr 2022 08:15:09	U	SMPL	
N2303-07 - 1	MW-103	-2	0.0144	-7470A	PPB	09 Apr 2022 08:17:24	U	SMPL	
N2303-08 - 1	MW-102	-16	0.0103	-7470A	PPB	09 Apr 2022 08:19:40	U	SMPL	
N2306-09 - 1	2431	2514	0.7494	-7470A	PPB	09 Apr 2022 08:21:55	U	SMPL	
N2306-10 - 1	2433	198	0.0728	-7470A	PPB	09 Apr 2022 08:24:10	U	SMPL	
N2306-12 - 1	2421-22	832	0.2580	-7470A	PPB	09 Apr 2022 08:26:26	U	SMPL	
N2306-14 - 1	2383-2386	148	0.0582	-7470A	PPB	09 Apr 2022 08:28:42	U	SMPL	
N2306-14DUP - 1	2383-2386DUP	114	0.0483	-7470A	PPB	09 Apr 2022 08:30:58	U	SMPL	
N2306-14MS - 1	2383-2386MS	12378	3.6309	-7470A	PPB	09 Apr 2022 08:33:32	U	SMPL	
N2306-14MSD - 1	2383-2386MSD	12390	3.6344	-7470A	PPB	09 Apr 2022 08:40:49	U	SMPL	
PB143985BL - 1	PBW	6	0.0167	-7470A	PPB	09 Apr 2022 08:43:04	U	SMPL	
CCV16 - 1	CCV16	16336	4.7872	-7470A	PPB	09 Apr 2022 08:45:30	U	SMPL	
CCB16 - 1	CCB16	-87	-0.0104	-7470A	PPB	09 Apr 2022 08:47:45	U	SMPL	
PB143985BS - 1	LCSW	15275	4.4772	-7470A	PPB	09 Apr 2022 08:50:02	U	SMPL	
N2292-01 - 1	2291	168	0.0641	-7470A	PPB	09 Apr 2022 08:52:18	U	SMPL	
N2292-02 - 1	2286	-39	0.0036	-7470A	PPB	09 Apr 2022 08:54:33	U	SMPL	
N2292-03 - 1	2287	14	0.0191	-7470A	PPB	09 Apr 2022 08:56:49	U	SMPL	
N2314-01 - 1	149TH-ST-STATION-DRUM-4-PPE	44	0.0278	-7470A	PPB	09 Apr 2022 08:59:05	U	SMPL	
N2314-02 - 1	149TH-ST-STATION-DRUM-5-POLY	47	0.0287	-7470A	PPB	09 Apr 2022 09:01:24	U	SMPL	
N2314-03 - 1	TREMONT-AVE-STATION-DRUM-14-PPE	38	0.0261	-7470A	PPB	09 Apr 2022 09:03:41	U	SMPL	
N2314-04 - 1	TREMONT-AVE-STATION-DRUM-15-POLY	54	0.0308	-7470A	PPB	09 Apr 2022 09:05:58	U	SMPL	
PB143986BL - 1	PBW	63	0.0334	-7470A	PPB	09 Apr 2022 09:08:15	U	SMPL	
PB143986BS - 1	LCSW	14984	4.3922	-7470A	PPB	09 Apr 2022 09:10:33	U	SMPL	
CCV17 - 1	CCV17	16768	4.9134	-7470A	PPB	09 Apr 2022 09:12:49	U	SMPL	
CCB17 - 1	CCB17	-100	-0.0142	-7470A	PPB	09 Apr 2022 09:15:10	U	SMPL	
N2288-01 - 1	TOTE	26	0.0226	-7470A	PPB	09 Apr 2022 09:17:27	U	SMPL	
N2288-01DUP - 1	TOTEDUP	47	0.0287	-7470A	PPB	09 Apr 2022 09:19:45	U	SMPL	
N2288-01MS - 1	TOTEMS	14330	4.2012	-7470A	PPB	09 Apr 2022 09:22:00	U	SMPL	
N2288-01MSD - 1	TOTEMSD	14058	4.1217	-7470A	PPB	09 Apr 2022 09:24:16	U	SMPL	
PB143898TB - 1	PB143898TB	-57	-0.0017	-7470A	PPB	09 Apr 2022 09:26:32	U	SMPL	
PB143956TB - 1	PB143956TB	13	0.0188	-7470A	PPB	09 Apr 2022 09:28:48	U	SMPL	
PB143982TB - 1	PB143982TB	35	0.0252	-7470A	PPB	09 Apr 2022 09:31:04	U	SMPL	
CCV18 - 1	CCV18	15740	4.6131	-7470A	PPB	09 Apr 2022 09:33:22	U	SMPL	
CCB18 - 1	CCB18	-143	-0.0268	-7470A	PPB	09 Apr 2022 09:35:39	U	SMPL	
N2272-08LX5 - 1		-23	0.0083	-7470A	PPB	09 Apr 2022 09:37:58	U	SMPL	
N2272-08A - 1		14702	4.3098	-7470A	PPB	09 Apr 2022 09:40:14	U	SMPL	
N2306-14LX5 - 1		-49	0.0007	-7470A	PPB	09 Apr 2022 09:42:30	U	SMPL	
N2306-14A - 1		14371	4.2131	-7470A	PPB	09 Apr 2022 09:44:53	U	SMPL	
N2288-01LX5 - 1		-87	-0.0104	-7470A	PPB	09 Apr 2022 09:47:11	U	SMPL	
N2288-01A - 1		14095	4.1325	-7470A	PPB	09 Apr 2022 09:49:28	U	SMPL	
CCV19 - 1	CCV19	16460	4.8234	-7470A	PPB	09 Apr 2022 09:54:18	U	SMPL	
CCB19 - 1	CCB19	-87	-0.0104	-7470A	PPB	09 Apr 2022 09:56:34	U	SMPL	