

LB 119604


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

31794



A= 0.0000e+000

B= 3.1305e-004 slope

C= 2.6733e-002 y-intercept

Rho= 0.9999035

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.034	0.034	23	0.000	23					=
0.05	0.050					185	0				15
0.20	0.200	0.229	0.029	646	0.0 %	646					-3
2.50	2.500	2.415	-0.085	7628	0.0 %	7628					-1
5.00	5.000	4.970	-0.030	15792	0.0 %	15792					1
7.50	7.500	7.572	0.072	24103	0.0 %	24103					0
10.0	10.000	9.980	-0.020	31794	0.0 %	31794					

LB119604
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	23	-		0 SFAM01.1	PPB	4/9/2022 10:24	Std
	0.2 S01	646	-		0.2 SFAM01.1	PPB	4/9/2022 10:26	Std
	2.5 S02	7628	-		2.5 SFAM01.1	PPB	4/9/2022 10:29	Std
	5 S03	15792	-		5 SFAM01.1	PPB	4/9/2022 10:31	Std
	7.5 S04	24103	-		7.5 SFAM01.1	PPB	4/9/2022 10:36	Std
	10 S05	31794	-		10 SFAM01.1	PPB	4/9/2022 10:41	Std
ICV069	ICV069	11814	3.7251	-	SFAM01.1	PPB	4/9/2022 10:44	SMPL
ICB069	ICB069	-59	0.0083	-	SFAM01.1	PPB	4/9/2022 10:46	SMPL
CCV020	CCV020	16323	5.1366	-	SFAM01.1	PPB	4/9/2022 10:48	SMPL
CCB020	CCB020	-54	0.0098	-	SFAM01.1	PPB	4/9/2022 10:50	SMPL
PB144005BL	PBW005	92	0.0555	-	SFAM01.1	PPB	4/9/2022 10:53	SMPL
N2163-01	YAYZ7	64	0.0468	-	SFAM01.1	PPB	4/9/2022 10:55	SMPL
N2163-02	YAZ25	94	0.0562	-	SFAM01.1	PPB	4/9/2022 10:57	SMPL
N2163-03	YAZ27	107	0.0602	-	SFAM01.1	PPB	4/9/2022 10:59	SMPL
N2163-04	YAZ29	137	0.0696	-	SFAM01.1	PPB	4/9/2022 11:02	SMPL
N2163-05	YAZ29D	119	0.064	-	SFAM01.1	PPB	4/9/2022 11:04	SMPL
N2163-06	YAZ29S	3902	1.2483	-	SFAM01.1	PPB	4/9/2022 11:11	SMPL
N2163-07	YAYY9	52	0.043	-	SFAM01.1	PPB	4/9/2022 11:13	SMPL
N2163-08	YAYZ1	109	0.0609	-	SFAM01.1	PPB	4/9/2022 11:15	SMPL
N2163-09	YAYZ3	109	0.0609	-	SFAM01.1	PPB	4/9/2022 11:18	SMPL
N2163-10	YAZ31	-2	0.0261	-	SFAM01.1	PPB	4/9/2022 11:20	SMPL
N2163-11	YAZ33	7	0.0289	-	SFAM01.1	PPB	4/9/2022 11:22	SMPL
N2163-12	YAYY5	984	0.3348	-	SFAM01.1	PPB	4/9/2022 11:24	SMPL
N2163-13	YAYZ5	986	0.3354	-	SFAM01.1	PPB	4/9/2022 11:27	SMPL
N2163-14	YAZ35	-10	0.0236	-	SFAM01.1	PPB	4/9/2022 11:29	SMPL
PB144006BL	PBW006	22	0.0336	-	SFAM01.1	PPB	4/9/2022 11:31	SMPL
N2164-01	YAYZ8	-14	0.0224	-	SFAM01.1	PPB	4/9/2022 11:33	SMPL
N2164-02	YAZ26	10	0.0299	-	SFAM01.1	PPB	4/9/2022 11:36	SMPL
N2164-03	YAZ28	-1	0.0264	-	SFAM01.1	PPB	4/9/2022 11:38	SMPL
N2164-04	YAZ30	65	0.0471	-	SFAM01.1	PPB	4/9/2022 11:40	SMPL
CCV021	CCV021	16199	5.0978	-	SFAM01.1	PPB	4/9/2022 11:42	SMPL
CCB021	CCB021	-27	0.0183	-	SFAM01.1	PPB	4/9/2022 11:45	SMPL
N2164-05	YAZ30D	76	0.0505	-	SFAM01.1	PPB	4/9/2022 11:47	SMPL
N2164-07	YAYZ0	67	0.0477	-	SFAM01.1	PPB	4/9/2022 11:51	SMPL
N2164-08	YAYZ2	69	0.0483	-	SFAM01.1	PPB	4/9/2022 11:54	SMPL
N2164-09	YAYZ4	27	0.0352	-	SFAM01.1	PPB	4/9/2022 11:56	SMPL
N2164-10	YAZ32	59	0.0452	-	SFAM01.1	PPB	4/9/2022 11:58	SMPL
N2164-11	YAZ34	29	0.0358	-	SFAM01.1	PPB	4/9/2022 12:01	SMPL
N2164-12	YAYY6	5747	1.8258	-	SFAM01.1	PPB	4/9/2022 12:03	SMPL
N2164-13	YAYZ6	2974	0.9577	-	SFAM01.1	PPB	4/9/2022 12:05	SMPL
N2164-14	YAZ36	28	0.0355	-	SFAM01.1	PPB	4/9/2022 12:07	SMPL
PB144007BL	PBW007	33	0.0371	-	SFAM01.1	PPB	4/9/2022 12:10	SMPL
N2188-01	MESQN3	72	0.0493	-	SFAM01.1	PPB	4/9/2022 12:12	SMPL
N2188-02	MESQQ2	-4	0.0255	-	SFAM01.1	PPB	4/9/2022 12:14	SMPL
N2188-03	MESQQ3	-3	0.0258	-	SFAM01.1	PPB	4/9/2022 12:16	SMPL
N2188-04	MESQQ4	6	0.0286	-	SFAM01.1	PPB	4/9/2022 12:19	SMPL

LB119604
INSTRUMENT ID : CV1

N2188-05	MESQQ4D	-1	0.0264	-	SFAM01.1	PPB	4/9/2022 12:21	SMPL
N2188-06	MESQQ4S	3890	1.2445	-	SFAM01.1	PPB	4/9/2022 12:23	SMPL
N2188-07	MESQR2	-30	0.0173	-	SFAM01.1	PPB	4/9/2022 12:26	SMPL
N2164-06	YAZ30S	3618	1.1593	-	SFAM01.1	PPB	4/9/2022 12:28	SMPL
CCV022	CCV022	16206	5.1	-	SFAM01.1	PPB	4/9/2022 12:30	SMPL
CCB022	CCB022	-31	0.017	-	SFAM01.1	PPB	4/9/2022 12:32	SMPL