


LB119609

7470A

INSTRUMENT ID: C112



A= 0.0000e+000

B= 3.0335e-004

C= 3.7928e-002

Rho= 0.9990926

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.071	0.071	108	0.000	108					
0.2	0.200	0.207	0.007	558	0.0 %	558					3
2.5	2.500	2.589	0.089	8409	0.0 %	8409					4
5.0	5.000	4.665	-0.335	15254	0.0 %	15254					-7
7.5	7.500	7.651	0.151	25098	0.0 %	25098					2
10.0	10.000	10.017	0.017	32895	0.0 %	32895					0

LB119609 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Apr 2022 08:54:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	108	-	0.0000	7470A	PPB	10 Apr 2022 08:58:08	S	Std
0.2 - 1	50.2	558	-	0.2000	7470A	PPB	10 Apr 2022 09:00:24	S	Std
2.5 - 1	52.5	8409	-	2.5000	7470A	PPB	10 Apr 2022 09:02:41	S	Std
5.0 - 1	55	15254	-	5.0000	7470A	PPB	10 Apr 2022 09:04:58	S	Std
7.5 - 1	57.5	25098	-	7.5000	7470A	PPB	10 Apr 2022 09:07:15	S	Std
10.0 - 1	510	32895	-	10.0000	7470A	PPB	10 Apr 2022 09:12:03	S	Std
ICV70 - 1	ICV70	13457	4.1201	-	7470A	PPB	10 Apr 2022 09:14:44	U	SMPL
ICB70 - 1	ICB70	-48	0.0234	-	7470A	PPB	10 Apr 2022 09:16:59	U	SMPL
CCV23 - 1	CCV23	16208	4.9546	-	7470A	PPB	10 Apr 2022 09:19:13	U	SMPL
CCB23 - 1	CCB23	-53	0.0219	-	7470A	PPB	10 Apr 2022 09:21:28	U	SMPL
CRA - 1	CRA	548	0.2042	-	7470A	PPB	10 Apr 2022 09:23:43	U	SMPL
HighStd - 1	HighStd	32920	10.0242	-	7470A	PPB	10 Apr 2022 09:25:57	U	SMPL
ChkStd - 1	ChkStd	23703	7.2282	-	7470A	PPB	10 Apr 2022 09:28:13	U	SMPL
PB144009BL - 1	PBW	-75	0.0152	-	7470A	PPB	10 Apr 2022 09:30:29	U	SMPL
N2285-02 - 1	COMP	412	0.1629	-	7470A	PPB	10 Apr 2022 09:35:02	U	SMPL
N2307-01 - 1	F05421	838	0.2921	-	7470A	PPB	10 Apr 2022 09:37:18	U	SMPL
N2327-01 - 1	A4128	76	0.0610	-	7470A	PPB	10 Apr 2022 09:39:35	U	SMPL
N2341-02 - 1	TT-040-IDWGW-20220407	181	0.0928	-	7470A	PPB	10 Apr 2022 09:41:52	U	SMPL
N2341-02DUP - 1	TT-040-IDWGW-20220407DUP	162	0.0871	-	7470A	PPB	10 Apr 2022 09:44:07	U	SMPL
CCV24 - 1	CCV24	15906	4.8630	-	7470A	PPB	10 Apr 2022 09:47:47	U	SMPL
CCB24 - 1	CCB24	-12	0.0343	-	7470A	PPB	10 Apr 2022 09:50:02	U	SMPL
N2341-02MS - 1	TT-040-IDWGW-20220407MS	12793	3.9187	-	7470A	PPB	10 Apr 2022 09:54:28	U	SMPL
N2341-02MSD - 1	TT-040-IDWGW-20220407MSD	13135	4.0224	-	7470A	PPB	10 Apr 2022 09:56:43	U	SMPL
PB144009BS - 1	LCSW	13863	4.2433	-	7470A	PPB	10 Apr 2022 09:58:58	U	SMPL
N2341-02LX5 - 1		61	0.0564	-	7470A	PPB	10 Apr 2022 10:01:14	U	SMPL
N2341-02A - 1		13121	4.0182	-	7470A	PPB	10 Apr 2022 10:03:29	U	SMPL
CCV25 - 1	CCV25	16190	4.9492	-	7470A	PPB	10 Apr 2022 10:05:45	U	SMPL
CCB25 - 1	CCB25	-50	0.0228	-	7470A	PPB	10 Apr 2022 10:08:01	U	SMPL