

LB119611


SFAM01.1

INSTRUMENT ID: CVA

Linear

 μ Abs.:

32164



A= 0.0000e+000

B= 3.1405e-004

C= -4.8882e-002

Rho= 0.9998624

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.004	0.004	167	0.000	167					
0.05	0.050					185	0				
0.20	0.200	0.226	0.026	876	0.0 %	876					
2.50	2.500	2.450	-0.050	7958	0.0 %	7958					
5.00	5.000	5.073	0.073	16310	0.0 %	16310					
7.50	7.500	7.394	-0.106	23700	0.0 %	23700					
10.0	10.000	10.052	0.052	32164	0.0 %	32164					

1
13
-2
-1
1

LB119611
INSTRUMENT ID :CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	167	-		0 SFAM01.1	PPB	4/10/2022 15:22	Std
	0.2 S01	876	-		0.2 SFAM01.1	PPB	4/10/2022 15:24	Std
	2.5 S02	7958	-		2.5 SFAM01.1	PPB	4/10/2022 15:28	Std
	5 S03	16310	-		5 SFAM01.1	PPB	4/10/2022 15:31	Std
	7.5 S04	23700	-		7.5 SFAM01.1	PPB	4/10/2022 15:33	Std
	10 S05	32164	-		10 SFAM01.1	PPB	4/10/2022 15:35	Std
ICV072	ICV072	12381	3.8394 -		SFAM01.1	PPB	4/10/2022 15:38	SMPL
ICB072	ICB072	88	-0.0212 -		SFAM01.1	PPB	4/10/2022 15:40	SMPL
CCV030	CCV030	15655	4.8676 -		SFAM01.1	PPB	4/10/2022 15:43	SMPL
CCB030	CCB030	84	-0.0225 -		SFAM01.1	PPB	4/10/2022 15:45	SMPL
PB144010BL	PBS010	17	-0.0435 -		SFAM01.1	PPB	4/10/2022 15:47	SMPL
N2184-01	MESQN2	1986	0.5748 -		SFAM01.1	PPB	4/10/2022 15:49	SMPL
N2184-02	MESQN7	4915	1.4947 -		SFAM01.1	PPB	4/10/2022 15:52	SMPL
N2184-03	MESQN8	2609	0.7705 -		SFAM01.1	PPB	4/10/2022 15:54	SMPL
N2184-04	MESQN9	2370	0.6954 -		SFAM01.1	PPB	4/10/2022 15:56	SMPL
N2184-05	MESQP0	301	0.0456 -		SFAM01.1	PPB	4/10/2022 15:58	SMPL
N2184-06	MESQP1	4116	1.2438 -		SFAM01.1	PPB	4/10/2022 16:01	SMPL
N2184-07	MESQP2	6351	1.9457 -		SFAM01.1	PPB	4/10/2022 16:03	SMPL
N2184-08	MESQP3	3864	1.1646 -		SFAM01.1	PPB	4/10/2022 16:05	SMPL
N2184-09	MESQP4	3361	1.0067 -		SFAM01.1	PPB	4/10/2022 16:07	SMPL
N2184-10	MESQP5	919	0.2397 -		SFAM01.1	PPB	4/10/2022 16:10	SMPL
N2184-11	MESQP6	2124	0.6182 -		SFAM01.1	PPB	4/10/2022 16:12	SMPL
N2184-12	MESQP6D	2473	0.7278 -		SFAM01.1	PPB	4/10/2022 16:14	SMPL
N2184-13	MESQP6S	9003	2.7785 -		SFAM01.1	PPB	4/10/2022 16:16	SMPL
N2184-14	MESQP7	5539	1.6907 -		SFAM01.1	PPB	4/10/2022 16:19	SMPL
N2184-15	MESQP8	732	0.181 -		SFAM01.1	PPB	4/10/2022 16:21	SMPL
PB144011BL	PBS011	88	-0.0212 -		SFAM01.1	PPB	4/10/2022 16:23	SMPL
N2187-02	MESQQ1	1331	0.3691 -		SFAM01.1	PPB	4/10/2022 16:25	SMPL
N2187-03	MESQQ6	377	0.0695 -		SFAM01.1	PPB	4/10/2022 16:28	SMPL
N2187-04	MESQQ7	323	0.0526 -		SFAM01.1	PPB	4/10/2022 16:30	SMPL
CCV031	CCV031	16582	5.1588 -		SFAM01.1	PPB	4/10/2022 16:32	SMPL
CCB031	CCB031	109	-0.0147 -		SFAM01.1	PPB	4/10/2022 16:35	SMPL
N2187-05	MESQQ9	804	0.2036 -		SFAM01.1	PPB	4/10/2022 16:37	SMPL
N2187-06	MESQR0	1187	0.3239 -		SFAM01.1	PPB	4/10/2022 16:39	SMPL
N2187-07	MESQR1	1404	0.392 -		SFAM01.1	PPB	4/10/2022 16:41	SMPL
N2187-08	MESQR1D	1289	0.3559 -		SFAM01.1	PPB	4/10/2022 16:44	SMPL
N2187-09	MESQR1S	9982	3.086 -		SFAM01.1	PPB	4/10/2022 16:46	SMPL
CCV032	CCV032	16477	5.1258 -		SFAM01.1	PPB	4/10/2022 16:48	SMPL
CCB032	CCB032	110	-0.0143 -		SFAM01.1	PPB	4/10/2022 16:50	SMPL