



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119613

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.7	04/08/2022	15:01
2	N2292-01		1	9	100.00	>212.0	>212.0	04/08/2022	15:20
3	N2292-02		1	10	100.00	>212.0	>212.0	04/08/2022	15:36
4	N2292-03		1	10	100.00	>212.0	>212.0	04/08/2022	15:52
5	N2292-03DUP		1	9	100.00	>212.0	>212.0	04/08/2022	16:09

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point 040822 WorkList ID : 158470 Department : Wet-Chemistry Date : 04-07-2022 11:49:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/11/2022	Solid	N2292-019	Flash Point	Cool 4 deg C	PSEG03	O11	2291	04/06/2022	1010B
04/11/2022	Solid	N2292-020	Flash Point	Cool 4 deg C	PSEG03	O11	2286	04/06/2022	1010B
04/11/2022	Solid	N2292-037	Flash Point	Cool 4 deg C	PSEG03	O11	2287	04/06/2022	1010B

Date/Time 4/8/22 14:40
Raw Sample Received by: AL (WCH)
Raw Sample Relinquished by: AL (WCH)

Date/Time 4/8/22 15:45
Raw Sample Received by: CP (WCH)
Raw Sample Relinquished by: AL (WCH)