



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119615

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/08/2022	16:20
2	CAL2	1	Water	NA	NA	20.7	7.01	04/08/2022	16:21
3	CAL3	1	Water	NA	NA	20.7	10.03	04/08/2022	16:22
4	ICV	1	Water	NA	NA	20.6	6.99	04/08/2022	16:24
5	CCV1	1	Water	NA	NA	20.6	2.01	04/08/2022	16:25
6	N2335-02	1	Solid	20.03	20	20.1	8.01	04/08/2022	16:30
7	N2335-02DUP	1	Solid	20.02	20	20.2	8.02	04/08/2022	16:31
8	N2336-02	1	Solid	20.04	20	20.3	8.64	04/11/2022	16:32
9	N2337-02	1	Solid	20.05	20	20.2	7.42	04/08/2022	16:33
10	CCV2	1	Water	NA	NA	20.7	12.03	04/08/2022	16:50

WORKLIST(Hardcopy Internal Chain)

VB 119615

WorkList Name : PH N2337 WorkList ID : 158522 Department : Wet-Chemistry Date : 04-08-2022 14:18:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/08/2022	Solid	N2335-02	pH	Cool 4 deg C	JPCL01	O11	A350-CU	04/08/2022	9045D
04/08/2022	Solid	N2336-02	pH	Cool 4 deg C	JPCL01	O11	C150-CU	04/08/2022	9045D
04/08/2022	Solid	N2337-02	pH	Cool 4 deg C	JPCL01	O11	B250-CU	04/08/2022	9045D

Date/Time 04/08/22 14:20
 Raw Sample Received by: ZP (99)
 Raw Sample Relinquished by: ZP SM

Date/Time 04/08/22 18:00
 Raw Sample Received by: ZP SM
 Raw Sample Relinquished by: ZP Wc