

LB119618


SFAM01.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

28914



A= 0.0000e+000

B= 3.2669e-004 slope

C= -2.2028e-002 y-intercept

Rho= 0.9952229

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RD
0.00	0.000	-0.008	-0.008	43	0.000	43					=
0.05	0.050					185	0				-3
0.20	0.200	0.194	-0.006	661	0.0 %	661					-8
2.50	2.500	2.290	-0.210	7078	0.0 %	7078					4
5.00	5.000	5.197	0.197	15974	0.0 %	15974					8
7.50	7.500	8.103	0.603	24872	0.0 %	24872					-6
10.0	10.000	9.424	-0.576	28914	0.0 %	28914					

LB119618
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	43	-		0 SFAM01.1	PPB	4/11/2022 11:03	Std
	0.2 S01	661	-		0.2 SFAM01.1	PPB	4/11/2022 11:06	Std
	2.5 S02	7078	-		2.5 SFAM01.1	PPB	4/11/2022 11:08	Std
	5 S03	15974	-		5 SFAM01.1	PPB	4/11/2022 11:15	Std
	7.5 S04	24872	-		7.5 SFAM01.1	PPB	4/11/2022 11:18	Std
	10 S05	28914	-		10 SFAM01.1	PPB	4/11/2022 11:31	Std
ICV073	ICV073	11486	3.7303	-	SFAM01.1	PPB	4/11/2022 11:38	SMPL
ICB073	ICB073	-19	-0.0282	-	SFAM01.1	PPB	4/11/2022 11:40	SMPL
CCV033	CCV033	14754	4.798	-	SFAM01.1	PPB	4/11/2022 11:45	SMPL
CCB033	CCB033	-35	-0.0335	-	SFAM01.1	PPB	4/11/2022 11:47	SMPL
PB144016BL	PBW016	45	-0.0073	-	SFAM01.1	PPB	4/11/2022 11:50	SMPL
N2066-01	YAYW7	343	0.09	-	SFAM01.1	PPB	4/11/2022 11:52	SMPL
N2066-04	YAZ45	24	-0.0142	-	SFAM01.1	PPB	4/11/2022 11:54	SMPL
N2066-05	YAZ45D	63	-0.0014	-	SFAM01.1	PPB	4/11/2022 11:56	SMPL
N2066-06	YAZ45S	3272	1.0469	-	SFAM01.1	PPB	4/11/2022 11:59	SMPL
N2066-07	YAZ09	52	-0.005	-	SFAM01.1	PPB	4/11/2022 12:01	SMPL
N2066-08	YAZ11	74	0.0021	-	SFAM01.1	PPB	4/11/2022 12:03	SMPL
N2066-09	YAZ43	41	-0.0086	-	SFAM01.1	PPB	4/11/2022 12:05	SMPL
N2066-10	YAZ47	65	-0.0008	-	SFAM01.1	PPB	4/11/2022 12:08	SMPL
N2066-11	YAZ49	55	-0.0041	-	SFAM01.1	PPB	4/11/2022 12:10	SMPL
N2066-12	YAZ51	75	0.0025	-	SFAM01.1	PPB	4/11/2022 12:12	SMPL
N2066-13	YAYW9	78	0.0035	-	SFAM01.1	PPB	4/11/2022 12:14	SMPL
N2066-14	YAYX1	73	0.0018	-	SFAM01.1	PPB	4/11/2022 12:17	SMPL
N2066-15	YAZ62	78	0.0035	-	SFAM01.1	PPB	4/11/2022 12:19	SMPL
N2066-16	YAZ64	49	-0.006	-	SFAM01.1	PPB	4/11/2022 12:21	SMPL
N2066-17	YAZ39	109	0.0136	-	SFAM01.1	PPB	4/11/2022 12:24	SMPL
N2066-18	YAZ57	50	-0.0057	-	SFAM01.1	PPB	4/11/2022 12:26	SMPL
N2066-19	YAZ07	165	0.0319	-	SFAM01.1	PPB	4/11/2022 12:28	SMPL
N2066-20	YAZ05	49	-0.006	-	SFAM01.1	PPB	4/11/2022 12:30	SMPL
N2066-21	YAZ53	59	-0.0028	-	SFAM01.1	PPB	4/11/2022 12:33	SMPL
CCV034	CCV034	15019	4.8845	-	SFAM01.1	PPB	4/11/2022 12:35	SMPL
CCB034	CCB034	2	-0.0214	-	SFAM01.1	PPB	4/11/2022 12:37	SMPL
PB144017BL	PBW017	37	-0.0099	-	SFAM01.1	PPB	4/11/2022 12:39	SMPL
N2067-01	YAYW8	223	0.0508	-	SFAM01.1	PPB	4/11/2022 12:42	SMPL
N2067-02	YAZ46	33	-0.0112	-	SFAM01.1	PPB	4/11/2022 12:44	SMPL
N2067-03	YAZ46D	68	0.0002	-	SFAM01.1	PPB	4/11/2022 12:46	SMPL
N2067-04	YAZ46S	3494	1.1194	-	SFAM01.1	PPB	4/11/2022 12:48	SMPL
N2067-05	YAZ10	100	0.0106	-	SFAM01.1	PPB	4/11/2022 12:51	SMPL
N2067-06	YAZ12	84	0.0054	-	SFAM01.1	PPB	4/11/2022 12:53	SMPL
N2067-07	YAZ38	278	0.0688	-	SFAM01.1	PPB	4/11/2022 12:55	SMPL
N2067-08	YAZ44	121	0.0175	-	SFAM01.1	PPB	4/11/2022 12:58	SMPL
N2067-09	YAZ48	48	-0.0063	-	SFAM01.1	PPB	4/11/2022 13:00	SMPL
N2067-10	YAZ50	43	-0.008	-	SFAM01.1	PPB	4/11/2022 13:02	SMPL
N2067-11	YAZ52	45	-0.0073	-	SFAM01.1	PPB	4/11/2022 13:04	SMPL
N2067-12	YAYX0	45	-0.0073	-	SFAM01.1	PPB	4/11/2022 13:07	SMPL
N2067-13	YAYX2	45	-0.0073	-	SFAM01.1	PPB	4/11/2022 13:09	SMPL

LB119618
INSTRUMENT ID : CV1

N2067-14	YAZ61	33	-0.0112 -	SFAM01.1	PPB	4/11/2022 13:11	SMPL
N2067-15	YAZ63	101	0.011 -	SFAM01.1	PPB	4/11/2022 13:13	SMPL
N2067-16	YAZ40	77	0.0031 -	SFAM01.1	PPB	4/11/2022 13:16	SMPL
N2067-17	YAZ58	46	-0.007 -	SFAM01.1	PPB	4/11/2022 13:18	SMPL
N2067-18	YAZ08	89	0.007 -	SFAM01.1	PPB	4/11/2022 13:20	SMPL
N2067-19	YAZ06	26	-0.0135 -	SFAM01.1	PPB	4/11/2022 13:22	SMPL
CCV035	CCV035	14475	4.7068 -	SFAM01.1	PPB	4/11/2022 13:25	SMPL
CCB035	CCB035	15	-0.0171 -	SFAM01.1	PPB	4/11/2022 13:27	SMPL
N2067-20	YAZ54	47	-0.0067 -	SFAM01.1	PPB	4/11/2022 13:29	SMPL
CCV036	CCV036	14814	4.8176 -	SFAM01.1	PPB	4/11/2022 13:32	SMPL
CCB036	CCB036	-4	-0.0233 -	SFAM01.1	PPB	4/11/2022 13:34	SMPL