



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119625

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/11/2022	15:00
2	CAL2	1	Water	NA	NA	20.3	7.00	04/11/2022	15:01
3	CAL3	1	Water	NA	NA	20.3	10.03	04/11/2022	15:02
4	ICV	1	Water	NA	NA	20.4	6.98	04/11/2022	15:05
5	CCV1	1	Water	NA	NA	20.4	2.01	04/11/2022	15:10
6	N2366-01	1	Water	NA	NA	20.4	4.55	04/11/2022	15:11
7	N2366-01DUP	1	Water	NA	NA	20.5	4.56	04/11/2022	15:12
8	CCV2	1	Water	NA	NA	20.4	12.02	04/11/2022	15:15

WORKLIST(Hardcopy Internal Chain)

L13119626

WorkList Name : IGNI-041122 WorkList ID : 158562 Department : Wet-Chemistry Date : 04-11-2022 09:14:10

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
04/12/2022	Solid	N2335-03	Ignitability	Cool 4 deg C	JPCL01	O11	A350-CU	04/08/2022 1030
04/12/2022	Solid	N2336-03	Ignitability	Cool 4 deg C	JPCL01	O11	C150-CU	04/08/2022 1030
04/12/2022	Solid	N2337-03	Ignitability	Cool 4 deg C	JPCL01	O11	B250-CU	04/08/2022 1030
04/21/2022	Solid	N2340-02	Ignitability	Cool 4 deg C	LIRO01	N13	SB-04-COMP	04/07/2022 1030
04/21/2022	Solid	N2340-05	Ignitability	Cool 4 deg C	LIRO01	N13	SB-06-COMP	04/07/2022 1030
04/21/2022	Solid	N2340-07	Ignitability	Cool 4 deg C	LIRO01	N13	SB-07-COMP	04/07/2022 1030
04/21/2022	Solid	N2340-09	Ignitability	Cool 4 deg C	LIRO01	N13	SB-08-COMP	04/07/2022 1030
04/21/2022	Solid	N2340-10	Ignitability	Cool 4 deg C	LIRO01	N13	SB-005-COMP	04/07/2022 1030
04/15/2022	Solid	N2349-20	Ignitability	Cool 4 deg C	PENN02	N12	COM-1	04/08/2022 1030

Date/Time 4/11/22 10:35
 Raw Sample Received by: R/LC
 Raw Sample Relinquished by: [Signature]

Date/Time 4/11/22 12:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: R/LC