



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119633

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/11/2022	17:37
2	CAL2	1	Water	NA	NA	20.7	7.00	04/11/2022	17:38
3	CAL3	1	Water	NA	NA	20.6	10.02	04/11/2022	17:39
4	ICV	1	Water	NA	NA	20.6	6.98	04/11/2022	17:40
5	CCV1	1	Water	NA	NA	20.6	2.01	04/11/2022	17:42
6	N2367-01	1	Solid	20.02	20	20.6	7.27	04/11/2022	17:43
7	N2367-01DUP	1	Solid	20.03	20	20.7	7.28	04/11/2022	17:44
8	CCV2	1	Water	NA	NA	20.6	12.03	04/11/2022	17:45