


LB119634

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3372e-004

C= -1.3587e-002

Rho= 0.9998259

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/2/1
0.0	0.000	-0.005	-0.005	35	0.000	35					2
0.2	0.200	0.235	0.035	1064	0.0 %	1064					18
2.5	2.500	2.535	0.035	10905	0.0 %	10905					1
5.0	5.000	4.858	-0.142	20842	0.0 %	20842					-3
7.5	7.500	7.575	0.075	32470	0.0 %	32470					1
10.0	10.000	10.003	0.003	42856	0.0 %	42856					0

LB119634 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Apr 2022 09:49:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0000	7470A	PPB	12 Apr 2022 10:02:42	S	Std
0.2 - 1	50.2	1064	-	0.2000	7470A	PPB	12 Apr 2022 10:04:59	S	Std
2.5 - 1	52.5	10905	-	2.5000	7470A	PPB	12 Apr 2022 10:07:15	S	Std
5.0 - 1	55	20842	-	5.0000	7470A	PPB	12 Apr 2022 10:09:32	S	Std
7.5 - 1	57.5	32470	-	7.5000	7470A	PPB	12 Apr 2022 10:14:25	S	Std
10.0 - 1	50	42856	-	10.0000	7470A	PPB	12 Apr 2022 10:19:47	S	Std
ICV74 - 1	ICV74	16422	3.8245	-	7470A	PPB	12 Apr 2022 10:22:33	U	SMPL
ICB74 - 1	ICB74	-72	-0.0304	-	7470A	PPB	12 Apr 2022 10:24:49	U	SMPL
CCV37 - 1	CCV37	21104	4.9188	-	7470A	PPB	12 Apr 2022 10:27:04	U	SMPL
CCB37 - 1	CCB37	-81	-0.0325	-	7470A	PPB	12 Apr 2022 10:29:19	U	SMPL
CRA - 1	CRA	912	0.1996	-	7470A	PPB	12 Apr 2022 10:31:35	U	SMPL
HighStd - 1	HighStd	43178	10.0778	-	7470A	PPB	12 Apr 2022 10:33:49	U	SMPL
ChkStd - 1	ChkStd	29662	6.9189	-	7470A	PPB	12 Apr 2022 10:36:04	U	SMPL
PB144047BL - 1	PBW	-123	-0.0423	-	7470A	PPB	12 Apr 2022 10:38:21	U	SMPL
PB144047BS - 1	LCSW	16497	3.8420	-	7470A	PPB	12 Apr 2022 10:40:37	U	SMPL
N2257-14 - 1	TP-1	-28	-0.0201	-	7470A	PPB	12 Apr 2022 10:42:53	U	SMPL
N2257-14DUP - 1	TP-1DUP	7	-0.0120	-	7470A	PPB	12 Apr 2022 10:45:10	U	SMPL
N2257-14MS - 1	TP-1MS	17392	4.0512	-	7470A	PPB	12 Apr 2022 10:50:07	U	SMPL
N2257-14MSD - 1	TP-1MSD	17378	4.0479	-	7470A	PPB	12 Apr 2022 10:52:23	U	SMPL
N2257-15 - 1	TP-2	-68	-0.0295	-	7470A	PPB	12 Apr 2022 10:54:38	U	SMPL
CCV38 - 1	CCV38	21347	4.9755	-	7470A	PPB	12 Apr 2022 10:56:54	U	SMPL
CCB38 - 1	CCB38	-67	-0.0292	-	7470A	PPB	12 Apr 2022 10:59:16	U	SMPL
N2257-16 - 1	TP-3	19	-0.0091	-	7470A	PPB	12 Apr 2022 11:01:33	U	SMPL
N2257-17 - 1	TP-4	38	-0.0047	-	7470A	PPB	12 Apr 2022 11:03:52	U	SMPL
N2257-18 - 1	TP-5	35	-0.0054	-	7470A	PPB	12 Apr 2022 11:06:08	U	SMPL
N2257-19 - 1	TP-6	29	-0.0068	-	7470A	PPB	12 Apr 2022 11:08:23	U	SMPL
N2335-03 - 1	A350-CU	37	-0.0049	-	7470A	PPB	12 Apr 2022 11:10:39	U	SMPL
N2336-03 - 1	C150-CU	30	-0.0066	-	7470A	PPB	12 Apr 2022 11:12:57	U	SMPL
N2337-03 - 1	B250-CU	39	-0.0045	-	7470A	PPB	12 Apr 2022 11:15:13	U	SMPL
N2340-02 - 1	SB-04-COMP	28	-0.0070	-	7470A	PPB	12 Apr 2022 11:17:35	U	SMPL
N2340-05 - 1	SB-06-COMP	45	-0.0031	-	7470A	PPB	12 Apr 2022 11:19:51	U	SMPL
N2340-07 - 1	SB-07-COMP	30	-0.0066	-	7470A	PPB	12 Apr 2022 11:22:20	U	SMPL
CCV39 - 1	CCV39	21070	4.9108	-	7470A	PPB	12 Apr 2022 11:24:36	U	SMPL
CCB39 - 1	CCB39	-56	-0.0267	-	7470A	PPB	12 Apr 2022 11:26:59	U	SMPL
N2340-09 - 1	SB-08-COMP	2	-0.0131	-	7470A	PPB	12 Apr 2022 11:29:16	U	SMPL
N2340-10 - 1	SB-005-COMP	7	-0.0120	-	7470A	PPB	12 Apr 2022 11:31:31	U	SMPL
PB143991TB - 1	PB143991TB	30	-0.0066	-	7470A	PPB	12 Apr 2022 11:33:47	U	SMPL
N2257-14LX5 - 1		1	-0.0134	-	7470A	PPB	12 Apr 2022 11:36:04	U	SMPL
N2257-14A - 1		16451	3.8313	-	7470A	PPB	12 Apr 2022 11:38:27	U	SMPL
CCV40 - 1	CCV40	20994	4.8930	-	7470A	PPB	12 Apr 2022 11:40:52	U	SMPL
CCB40 - 1	CCB40	-57	-0.0269	-	7470A	PPB	12 Apr 2022 11:43:15	U	SMPL