


LB119645

SFAM01.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.014	-0.014	71	0.000	71					=
0.05	0.050					185	0				4
0.20	0.200	0.208	0.008	1051	0.0 %	1051					-1
2.50	2.500	2.463	-0.037	11011	0.0 %	11011					1
5.00	5.000	5.070	0.070	22529	0.0 %	22529					0
7.50	7.500	7.500	0.000	33261	0.0 %	33261					0
10.0	10.000	9.973	-0.027	44184	0.0 %	44184					0

LB119645
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	71	-		0 SFAM01.1	PPB	4/12/2022 14:11	Std
	0.2 S01	1051	-		0.2 SFAM01.1	PPB	4/12/2022 14:13	Std
	2.5 S02	11011	-		2.5 SFAM01.1	PPB	4/12/2022 14:16	Std
	5 S03	22529	-		5 SFAM01.1	PPB	4/12/2022 14:18	Std
	7.5 S04	33261	-		7.5 SFAM01.1	PPB	4/12/2022 14:20	Std
	10 S05	44184	-		10 SFAM01.1	PPB	4/12/2022 14:22	Std
ICV075	ICV075	16439	3.6916 -		SFAM01.1	PPB	4/12/2022 14:25	SMPL
ICB075	ICB075	-29	-0.0368 -		SFAM01.1	PPB	4/12/2022 14:28	SMPL
CCV041	CCV041	22031	4.9577 -		SFAM01.1	PPB	4/12/2022 14:30	SMPL
CCB041	CCB041	-34	-0.0379 -		SFAM01.1	PPB	4/12/2022 14:32	SMPL
PB144070BL	PBW070	41	-0.0209 -		SFAM01.1	PPB	4/12/2022 14:34	SMPL
N2228-01	YAY37	398006	HIGH -		SFAM01.1	PPB	4/12/2022 14:37	
N2228-02	YAY39	437937	HIGH -		SFAM01.1	PPB	4/12/2022 14:38	
N2228-03	YAY43	15410	3.4586 -		SFAM01.1	PPB	4/12/2022 14:40	SMPL
N2228-04	YAY45	15229	3.4177 -		SFAM01.1	PPB	4/12/2022 14:43	SMPL
N2228-05	YAY49	4510	0.9909 -		SFAM01.1	PPB	4/12/2022 14:45	SMPL
N2228-06	YAY51	4850	1.0678 -		SFAM01.1	PPB	4/12/2022 14:47	SMPL
N2228-07	YAY53	152663	34.5329 -		SFAM01.1	PPB	4/12/2022 14:49	SMPL
N2228-08	YAY57	10296	2.3008 -		SFAM01.1	PPB	4/12/2022 14:52	SMPL
N2228-09	YAY63	3460	0.7531 -		SFAM01.1	PPB	4/12/2022 14:54	SMPL
N2228-10	YAY65	3510	0.7645 -		SFAM01.1	PPB	4/12/2022 14:56	SMPL
N2228-11	YAY69	465	0.0751 -		SFAM01.1	PPB	4/12/2022 14:58	SMPL
N2228-12	YAY69D	526	0.0889 -		SFAM01.1	PPB	4/12/2022 15:01	SMPL
N2228-13	YAY69S	4713	1.0368 -		SFAM01.1	PPB	4/12/2022 15:07	SMPL
N2228-14	YAY87	16275	3.6545 -		SFAM01.1	PPB	4/12/2022 15:10	SMPL
N2228-15	YAY73	5172	1.1407 -		SFAM01.1	PPB	4/12/2022 15:12	SMPL
N2228-16	YAY75	6686	1.4835 -		SFAM01.1	PPB	4/12/2022 15:14	SMPL
N2228-17	YAY77	1272	0.2578 -		SFAM01.1	PPB	4/12/2022 15:17	SMPL
N2228-18	YAY81	306	0.0391 -		SFAM01.1	PPB	4/12/2022 15:19	SMPL
N2228-19	YAY89	802	0.1514 -		SFAM01.1	PPB	4/12/2022 15:21	SMPL
CCV042	CCV042	22965	5.1691 -		SFAM01.1	PPB	4/12/2022 15:23	SMPL
CCB042	CCB042	-15	-0.0336 -		SFAM01.1	PPB	4/12/2022 15:26	SMPL
N2228-20	YAY83	40	-0.0211 -		SFAM01.1	PPB	4/12/2022 15:28	SMPL
N2228-21	YAY85	1410	0.289 -		SFAM01.1	PPB	4/12/2022 15:30	SMPL
PB144071BL	PBW071	12	-0.0275 -		SFAM01.1	PPB	4/12/2022 15:32	SMPL
PB144072BL	PBW072	-57	-0.0431 -		SFAM01.1	PPB	4/12/2022 15:43	SMPL
N2231-01	YAY38	21047	4.7349 -		SFAM01.1	PPB	4/12/2022 15:45	SMPL
N2231-02	YAY40	3497	0.7615 -		SFAM01.1	PPB	4/12/2022 15:48	SMPL
N2231-03	YAY44	482	0.0789 -		SFAM01.1	PPB	4/12/2022 15:50	SMPL
N2231-04	YAY46	820	0.1554 -		SFAM01.1	PPB	4/12/2022 15:52	SMPL
N2231-05	YAY56	268	0.0305 -		SFAM01.1	PPB	4/12/2022 15:54	SMPL
N2231-06	YAY50	914	0.1767 -		SFAM01.1	PPB	4/12/2022 15:57	SMPL
N2231-07	YAY52	899	0.1733 -		SFAM01.1	PPB	4/12/2022 15:59	SMPL
N2231-08	YAY54	2870	0.6196 -		SFAM01.1	PPB	4/12/2022 16:01	SMPL
N2231-09	YAY58	904	0.1745 -		SFAM01.1	PPB	4/12/2022 16:04	SMPL
N2231-10	YAY91	137	0.0008 -		SFAM01.1	PPB	4/12/2022 16:06	SMPL

LB119645
INSTRUMENT ID : CV1

N2231-11	YAY64	426	0.0662 -	SFAM01.1	PPB	4/12/2022 16:08 SMPL
N2231-12	YAY66	362	0.0518 -	SFAM01.1	PPB	4/12/2022 16:10 SMPL
N2231-13	YAY70	229	0.0216 -	SFAM01.1	PPB	4/12/2022 16:13 SMPL
CCV043	CCV043	22466	5.0561 -	SFAM01.1	PPB	4/12/2022 16:15 SMPL
CCB043	CCB043	0	-0.0302 -	SFAM01.1	PPB	4/12/2022 16:17 SMPL
N2231-14	YAY70D	208	0.0169 -	SFAM01.1	PPB	4/12/2022 16:19 SMPL
N2231-15	YAY70S	4709	1.0359 -	SFAM01.1	PPB	4/12/2022 16:24 SMPL
N2231-16	YAY88	145	0.0026 -	SFAM01.1	PPB	4/12/2022 16:26 SMPL
N2231-17	YAY74	320	0.0422 -	SFAM01.1	PPB	4/12/2022 16:29 SMPL
N2231-18	YAY76	356	0.0504 -	SFAM01.1	PPB	4/12/2022 16:31 SMPL
N2231-19	YAY78	235	0.023 -	SFAM01.1	PPB	4/12/2022 16:33 SMPL
N2231-20	YAY82	115	-0.0042 -	SFAM01.1	PPB	4/12/2022 16:35 SMPL
N2231-21	YAY90	366	0.0527 -	SFAM01.1	PPB	4/12/2022 16:38 SMPL
N2231-22	YAY84	67	-0.015 -	SFAM01.1	PPB	4/12/2022 16:40 SMPL
PB144073BL	PBW073	68	-0.0148 -	SFAM01.1	PPB	4/12/2022 16:42 SMPL
N2232-01	YAY62	97	-0.0082 -	SFAM01.1	PPB	4/12/2022 16:45 SMPL
N2232-02	YAY62D	109	-0.0055 -	SFAM01.1	PPB	4/12/2022 16:47 SMPL
N2232-03	YAY62S	4516	0.9922 -	SFAM01.1	PPB	4/12/2022 16:49 SMPL
N2232-04	YAY86	96	-0.0085 -	SFAM01.1	PPB	4/12/2022 16:51 SMPL
N2228-01DLX10	YAY37	67076	15.1559 -	SFAM01.1	PPB	4/12/2022 16:54 SMPL
N2228-02DLX10	YAY39	72708	16.431 -	SFAM01.1	PPB	4/12/2022 16:56 SMPL
N2228-07DLX10	YAY53	15200	3.4111 -	SFAM01.1	PPB	4/12/2022 16:58 SMPL
N2230-01	YAY61	21431	4.8218 -	SFAM01.1	PPB	4/12/2022 17:01 SMPL
N2230-02	YAY61D	21730	4.8895 -	SFAM01.1	PPB	4/12/2022 17:03 SMPL
N2230-03	YAY61S	25852	5.8227 -	SFAM01.1	PPB	4/12/2022 17:06 SMPL
CCV044	CCV044	22940	5.1635 -	SFAM01.1	PPB	4/12/2022 17:09 SMPL
CCB044	CCB044	-28	-0.0365 -	SFAM01.1	PPB	4/12/2022 17:11 SMPL
N2228-01DL2X50	YAY37	13008	2.9148 -	SFAM01.1	PPB	4/12/2022 17:13 SMPL
N2228-02DL2X50	YAY39	14755	3.3104 -	SFAM01.1	PPB	4/12/2022 17:15 SMPL
CCV045	CCV045	23085	5.1963 -	SFAM01.1	PPB	4/12/2022 17:18 SMPL
CCB045	CCB045	-60	-0.0438 -	SFAM01.1	PPB	4/12/2022 17:20 SMPL